

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
 Data File : BF133929.D
 Acq On : 23 Jun 2023 06:18
 Operator : CG\JU
 Sample : 03320-04
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-10

Quant Time: Jun 23 06:38:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 02:21:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	106463	20.000	ng	0.00
21) Naphthalene-d8	8.139	136	387417	20.000	ng	0.00
39) Acenaphthene-d10	9.898	164	175560	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	290596	20.000	ng	0.00
76) Chrysene-d12	14.051	240	210836	20.000	ng	0.00
86) Perylene-d12	15.557	264	155880	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	608502	99.503	ng	0.00
7) Phenol-d6	6.498	99	740977	93.085	ng	0.00
23) Nitrobenzene-d5	7.422	82	504069	70.867	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	185190	83.333	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	894549	72.432	ng	0.00
79) Terphenyl-d14	12.986	244	815952	59.883	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
Data File : BF133929.D
Acq On : 23 Jun 2023 06:18
Operator : CG\JU
Sample : 03320-04
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-10

Quant Time: Jun 23 06:38:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 02:21:26 2023
Response via : Initial Calibration

