


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
Data File : BF133888.D
Acq On : 22 Jun 2023 07:58
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Jun 23 02:21:26 2023



AutoFind: Scans 1624, 1625, 1626; Background Corrected with Scan 1619

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.9	63536	PASS
68	69	0.00	2	1.9	1013	PASS
69	198	0.00	100	38.0	53757	PASS
70	69	0.00	2	0.6	321	PASS
127	198	10	80	47.6	67373	PASS
197	198	0.00	2	0.5	761	PASS
198	198	100	100	100.0	141403	PASS
199	198	5	9	6.9	9793	PASS
275	198	10	60	28.0	39597	PASS
365	198	1	100	3.5	4955	PASS
441	198	0.01	100	14.0	19817	PASS
442	442	50	100	100.0	122664	PASS
443	442	15	24	18.8	23032	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
6/22/2023	BNA_F	<u>BF133888.D</u>
Compound Name	Response	Retention Time
DDT	324776	13.568
DDD	12631	13.268
DDE	266	12.927
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12897	337673	3.82

Instrument :
BNA_F
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
 Data File : BF133888.D
 Acq On : 22 Jun 2023 07:58
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jun 23 04:15:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 02:21:26 2023
 Response via : Initial Calibration



TIC: BF133888.D\data.ms

(70) Pentachlorophenol (C)

11.198min (+ 0.000) 34271.36 ng




response 139943

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	69.60	67.67
264.00	63.40	60.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
Data File : BF133888.D
Acq On : 22 Jun 2023 07:58
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jun 23 04:15:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 02:21:26 2023
Response via : Initial Calibration



TIC: BF133888.D\data.ms

(77) Benzidine

12.739min (+ 0.000) 66135.69 ng

response 645801

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.36
183.00	10.50	11.07
0.00	0.00	0.00