

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF062323\  
 Data File : BF133932.D  
 Acq On : 23 Jun 2023 09:12  
 Operator : CG\JU  
 Sample : PB153615BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB153615BL

Quant Time: Jun 25 02:34:54 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF060923.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jun 23 02:21:26 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.857  | 152  | 103625   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.139  | 136  | 407823   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.898  | 164  | 209636   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.392 | 188  | 430091   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 14.051 | 240  | 250446   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.562 | 264  | 212927   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.504  | 112  | 982083   | 164.990 | ng    | 0.01     |
| 7) Phenol-d6                | 6.504  | 99   | 1189026  | 153.462 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.428  | 82   | 818292   | 109.287 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.698 | 330  | 411846   | 155.202 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.222  | 172  | 1484377  | 100.654 | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.998 | 244  | 1808992  | 111.766 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF062323\  
Data File : BF133932.D  
Acq On : 23 Jun 2023 09:12  
Operator : CG\JU  
Sample : PB153615BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB153615BL

Quant Time: Jun 25 02:34:54 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF060923.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jun 23 02:21:26 2023  
Response via : Initial Calibration

