

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062618\
 Data File : BF106832.D
 Acq On : 26 Jun 2018 19:27
 Operator : JU/SJ
 Sample : J3642-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEST-PIT-2

Manual Integrations
 APPROVED

Sohil
 6/27/2018 4:18:30 PM

Quant Time: Jun 27 09:29:35 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	42337	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	145284	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	63130	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	125991	20.00	ng	0.00
75) Chrysene-d12	14.15	240	127469	20.00	ng	-0.01
86) Perylene-d12	15.69	264	86599	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	251749	100.69	ng	0.01
7) Phenol-d6	6.61	99	304023	97.37	ng	0.00
23) Nitrobenzene-d5	7.52	82	187161	71.59	ng	-0.01
41) 2,4,6-Tribromophenol	10.80	330	85172	116.40	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	325725	87.82	ng	0.00
78) Terphenyl-d14	13.08	244	386907	73.52	ng	0.00

Target Compounds				Qvalue		
20) 3+4-Methylphenols	7.38	107	234	Below Cal	#	20
31) Naphthalene	8.27	128	30013	4.419	ng	97
37) 2-Methylnaphthalene	8.95	142	12584	2.795	ng	94
48) Acenaphthylene	9.87	152	29214	4.672	ng	98
49) Dimethylphthalate	9.71	163	58134	12.278	ng	98
57) Fluorene	10.55	166	9905	2.315	ng	# 88
70) Phenanthrene	11.52	178	120400	19.813	ng	98
71) Anthracene	11.57	178	38928	6.276	ng	99
74) Fluoranthene	12.72	202	276848	41.334	ng	94
77) Pyrene	12.95	202	236447	28.129	ng	100
80) Benzo(a)anthracene	14.14	228	79844	10.277	ng	92
82) Chrysene	14.18	228	115437	16.151	ng	98
83) Bis(2-ethylhexyl)phthalate	14.10	149	13851	3.135	ng	# 91
85) Indeno(1,2,3-cd)pyrene	17.27	276	40467	6.672	ng	# 84
87) Benzo(b)fluoranthene	15.24	252	94228m	17.516	ng	
88) Benzo(k)fluoranthene	15.27	252	34015m	6.640	ng	
89) Benzo(a)pyrene	15.63	252	54405	11.376	ng	94
90) Dibenzo(a,h)anthracene	17.28	278	9664	2.384	ng	# 87
91) Benzo(a,h,i)perylene	17.75	276	47853	12.080	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062618\
Data File : BF106832.D
Acq On : 26 Jun 2018 19:27
Operator : JU/SJ
Sample : J3642-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEST-PIT-2

Manual Integrations
APPROVED

Sohil
6/27/2018 4:18:30 PM

Quant Time: Jun 27 09:29:35 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

