

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062618\
 Data File : BF106854.D
 Acq On : 27 Jun 2018 5:59
 Operator : JU/SJ
 Sample : J3639-07
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 062018-TIPC-E-D-M

Quant Time: Jun 27 06:47:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	42572	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	146977	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	64891	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	125926	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	127532	20.00	ng	-0.02
86) Perylene-d12	15.67	264	84749	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	93208	37.07	ng	0.00
7) Phenol-d6	6.60	99	72656	23.14	ng	-0.02
23) Nitrobenzene-d5	7.52	82	213040	80.55	ng	-0.01
41) 2,4,6-Tribromophenol	10.79	330	54440	72.38	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	396293	107.23	ng	-0.01
78) Terphenyl-d14	13.08	244	453981	86.23	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.71	163	14685	3.017	ng	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062618\
Data File : BF106854.D
Acq On : 27 Jun 2018 5:59
Operator : JU/SJ
Sample : J3639-07
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
062018-TIPC-E-D-M

Quant Time: Jun 27 06:47:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

