


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062618\
 Data File : BF106839.D
 Acq On : 26 Jun 2018 23:24
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF061918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 25 18:50:17 2018



AutoFind: Scans 1595, 1596, 1597; Background Corrected with Scan 1591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.5	5895	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	28.5	5703	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	37.9	7578	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	20003	PASS
199	198	5	9	6.7	1345	PASS
275	198	10	60	27.3	5457	PASS
365	198	1	100	2.5	494	PASS
441	443	0.01	100	81.0	2862	PASS
442	198	50	100	90.0	17999	PASS
443	442	15	24	19.6	3532	PASS