

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
 Data File : BF106867.D
 Acq On : 27 Jun 2018 13:12
 Operator : JU/SJ
 Sample : J3684-10
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleID :
 TP-1-D

Manual Integrations
 APPROVED

Sohil
 6/28/2018 2:21:30 PM

Quant Time: Jun 27 14:31:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	47352	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	184788	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	96553	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	195869	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	140996	20.00	ng	-0.02
86) Perylene-d12	15.70	264	144480	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	278437	99.57	ng	0.00
7) Phenol-d6	6.60	99	359312	102.89	ng	-0.02
23) Nitrobenzene-d5	7.52	82	225695	67.88	ng	-0.02
41) 2,4,6-Tribromophenol	10.80	330	127342	113.79	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	386402	64.11	ng	-0.02
78) Terphenyl-d14	13.07	244	431553	74.14	ng	-0.01
Target Compounds						
10) Phenol	6.61	94	8214	2.061	ng	90
49) Dimethylphthalate	9.70	163	86060	11.884	ng	98
51) Acenaphthene	10.02	154	13320	2.212	ng	97
57) Fluorene	10.54	166	21139	3.231	ng	97
70) Phenanthrene	11.51	178	238513	25.247	ng	99
71) Anthracene	11.57	178	67831	7.034	ng	99
74) Fluoranthene	12.71	202	370031	35.537	ng	93
77) Pyrene	12.94	202	331278	35.630	ng	99
80) Benzo(a)anthracene	14.14	228	162710	18.934	ng	99
82) Chrysene	14.18	228	143410	18.140	ng	97
85) Indeno(1,2,3-cd)pyrene	17.28	276	73294	10.925	ng	# 87
87) Benzo(b)fluoranthene	15.25	252	180136m	20.071	ng	
88) Benzo(k)fluoranthene	15.27	252	47483m	5.556	ng	
89) Benzo(a)pyrene	15.57	252	77947	9.770	ng	98
90) Dibenzo(a,h)anthracene	17.28	278	19941	2.949	ng	# 70
91) Benzo(g,h,i)perylene	17.75	276	66260	10.026	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
Data File : BF106867.D
Acq On : 27 Jun 2018 13:12
Operator : JU/SJ
Sample : J3684-10
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
TP-1-D

Manual Integrations
APPROVED

Sohil
6/28/2018 2:21:30 PM

Quant Time: Jun 27 14:31:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

