

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
 Data File : BF106873.D
 Acq On : 27 Jun 2018 15:50
 Operator : JU/SJ
 Sample : J3570-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20180373-AMENDED-COMPOSITE-

Manual Integrations
 APPROVED

Sohil
 6/28/2018 2:21:33 PM

Quant Time: Jun 28 11:10:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	45012	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	176913	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	93607	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	173857	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	125756	20.00	ng	-0.02
86) Perylene-d12	15.69	264	138127	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	20155	7.58	ng	0.00
7) Phenol-d6	6.60	99	115694	34.85	ng	-0.02
23) Nitrobenzene-d5	7.52	82	194385	61.06	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	7882	7.26	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	351585	59.08	ng	-0.01
78) Terphenyl-d14	13.07	244	344501	66.36	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.70	163	47199	6.723	ng	# 98
74) Fluoranthene	12.71	202	23619	2.555	ng	96
77) Pyrene	12.94	202	32018	3.861	ng	97
83) Bis(2-ethylhexyl)phthalate	14.09	149	11388	2.613	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
Data File : BF106873.D
Acq On : 27 Jun 2018 15:50
Operator : JU/SJ
Sample : J3570-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20180373-AMENDED-COMPOSITE-

Manual Integrations
APPROVED

Sohil
6/28/2018 2:21:33 PM

Quant Time: Jun 28 11:10:33 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

