

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
 Data File : BF106888.D
 Acq On : 27 Jun 2018 22:26
 Operator : JU/SJ
 Sample : J3247-07
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OR-02-062618

Quant Time: Jun 28 04:53:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	42594	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	153331	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	66151	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	118314	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	118524	20.00	ng	-0.02
86) Perylene-d12	15.68	264	84906	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	163756	65.10	ng	0.00
7) Phenol-d6	6.60	99	201850	64.26	ng	-0.02
23) Nitrobenzene-d5	7.52	82	119093	43.16	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	49162	64.12	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	223435	51.33	ng	-0.02
78) Terphenyl-d14	13.07	244	243912	49.85	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.70	163	50726	10.224	ng	97
74) Fluoranthene	12.71	202	14806	2.354	ng	94
77) Pyrene	12.94	202	17695	2.264	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
Data File : BF106888.D
Acq On : 27 Jun 2018 22:26
Operator : JU/SJ
Sample : J3247-07
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OR-02-062618

Quant Time: Jun 28 04:53:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

