

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
 Data File : BF106898.D
 Acq On : 28 Jun 2018 3:48
 Operator : JU/SJ
 Sample : J3705-04
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 W-3

Quant Time: Jun 28 07:13:49 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	46093	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	161494	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	68358	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	118159	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	130597	20.00	ng	-0.02
86) Perylene-d12	15.68	264	92509	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	156383	57.45	ng	0.00
7) Phenol-d6	6.60	99	115216	33.89	ng	-0.02
23) Nitrobenzene-d5	7.52	82	264952	91.18	ng	-0.01
41) 2,4,6-Tribromophenol	10.79	330	103144	130.19	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	443943	115.16	ng	-0.01
78) Terphenyl-d14	13.08	244	490584	90.99	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.70	163	12033	2.347	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062718\
Data File : BF106898.D
Acq On : 28 Jun 2018 3:48
Operator : JU/SJ
Sample : J3705-04
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
W-3

Quant Time: Jun 28 07:13:49 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

