

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062723\
 Data File : BF133993.D
 Acq On : 27 Jun 2023 16:11
 Operator : CG\JU
 Sample : 03337-07
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-9

Quant Time: Jun 27 16:46:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 26 23:12:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.851	152	109252	20.000	ng	0.00
21) Naphthalene-d8	8.128	136	398890	20.000	ng	0.00
39) Acenaphthene-d10	9.886	164	182689	20.000	ng	0.00
64) Phenanthrene-d10	11.380	188	323248	20.000	ng	0.00
76) Chrysene-d12	14.039	240	193630	20.000	ng	0.00
86) Perylene-d12	15.545	264	187859	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	550471	84.283	ng	0.00
7) Phenol-d6	6.486	99	701940	87.392	ng	0.00
23) Nitrobenzene-d5	7.410	82	416820	56.328	ng	0.00
42) 2,4,6-Tribromophenol	10.680	330	173214	75.621	ng	0.00
45) 2-Fluorobiphenyl	9.210	172	841995	67.008	ng	0.00
79) Terphenyl-d14	12.980	244	678824	56.423	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062723\
Data File : BF133993.D
Acq On : 27 Jun 2023 16:11
Operator : CG\JU
Sample : 03337-07
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-9

Quant Time: Jun 27 16:46:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 26 23:12:04 2023
Response via : Initial Calibration

