

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF062723\  
 Data File : BF133996.D  
 Acq On : 27 Jun 2023 17:54  
 Operator : CG\JU  
 Sample : 03357-11  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 WC-SCAN-03

Quant Time: Jun 27 18:23:05 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF062623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jun 26 23:12:04 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 6.851  | 152  | 101959   | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8            | 8.128  | 136  | 372535   | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10          | 9.886  | 164  | 174695   | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10          | 11.380 | 188  | 286526   | 20.000 | ng    | 0.00     |
| 76) Chrysene-d12              | 14.057 | 240  | 138420   | 20.000 | ng    | # 0.01   |
| 86) Perylene-d12              | 15.574 | 264  | 126308   | 20.000 | ng    | 0.03     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 5) 2-Fluorophenol             | 5.487  | 112  | 495267   | 81.255 | ng    | 0.00     |
| 7) Phenol-d6                  | 6.487  | 99   | 618855   | 82.559 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 7.410  | 82   | 419496   | 60.701 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 10.680 | 330  | 125183   | 57.153 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 9.204  | 172  | 824377   | 68.608 | ng    | 0.00     |
| 79) Terphenyl-d14             | 12.986 | 244  | 549527   | 63.894 | ng    | 0.00     |
| Target Compounds              |        |      |          |        |       |          |
| 84) Bis(2-ethylhexyl)phtha... | 14.033 | 149  | 34556    | 6.225  | ng    | # 80     |
| -----                         |        |      |          |        |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF062723\  
Data File : BF133996.D  
Acq On : 27 Jun 2023 17:54  
Operator : CG\JU  
Sample : 03357-11  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
WC-SCAN-03

Quant Time: Jun 27 18:23:05 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF062623.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 26 23:12:04 2023  
Response via : Initial Calibration

