

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062818\
 Data File : BF106919.D
 Acq On : 28 Jun 2018 14:36
 Operator : JU/SJ
 Sample : J3570-03RE
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20180373-AMENDED-COMPOSITE-

Manual Integrations
 APPROVED

Sohil
 6/29/2018 2:30:37 PM

Quant Time: Jun 28 17:31:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 25 18:50:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	79428	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	309783	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	158012	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	292160	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	206507	20.00	ng	-0.02
86) Perylene-d12	15.70	264	228377	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	36020	7.68	ng	0.00
7) Phenol-d6	6.59	99	201564	34.41	ng	-0.02
23) Nitrobenzene-d5	7.52	82	336144	60.30	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	13377	7.30	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	597236	59.56	ng	-0.02
78) Terphenyl-d14	13.07	244	552932	64.86	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.70	163	82186	6.935	ng	# 96
74) Fluoranthene	12.71	202	39337	2.533	ng	96
77) Pyrene	12.94	202	51685	3.795	ng	98
83) Bis(2-ethylhexyl)phthalate	14.10	149	17101	2.389	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062818\
Data File : BF106919.D
Acq On : 28 Jun 2018 14:36
Operator : JU/SJ
Sample : J3570-03RE
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20180373-AMENDED-COMPOSITE-

Manual Integrations
APPROVED

Sohil
6/29/2018 2:30:37 PM

Quant Time: Jun 28 17:31:03 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 25 18:50:17 2018
Response via : Initial Calibration

