


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF062818\  
 Data File : BF106951.D  
 Acq On : 29 Jun 2018 4:55  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF061918.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Jun 25 18:50:17 2018



AutoFind: Scans 1594, 1595, 1596; Background Corrected with Scan 1590

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 28.4      | 8136    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 28.2      | 8063    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 37.2      | 10637   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 28625   | PASS             |
| 199         | 198          | 5            | 9            | 7.1       | 2021    | PASS             |
| 275         | 198          | 10           | 60           | 28.4      | 8126    | PASS             |
| 365         | 198          | 1            | 100          | 3.5       | 992     | PASS             |
| 441         | 443          | 0.01         | 100          | 86.5      | 4455    | PASS             |
| 442         | 198          | 50           | 100          | 95.3      | 27271   | PASS             |
| 443         | 442          | 15           | 24           | 18.9      | 5151    | PASS             |