

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063023\
 Data File : BF134051.D
 Acq On : 30 Jun 2023 14:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 30 18:36:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	38.996	2.5	108	0.00
3	Pyridine	40.000	41.445	-3.6	115	0.00
4	n-Nitrosodimethylamine	40.000	41.430	-3.6	113	0.00
5 S	2-Fluorophenol	80.000	79.053	1.2	111	0.00
6	Aniline	40.000	39.437	1.4	109	0.00
7 S	Phenol-d6	80.000	77.376	3.3	110	0.00
8	2-Chlorophenol	40.000	39.120	2.2	110	0.00
9	Benzaldehyde	40.000	40.477	-1.2	111	0.00
10 C	Phenol	40.000	38.626	3.4	109	0.00
11	bis(2-Chloroethyl)ether	40.000	38.635	3.4	108	0.00
12	1,3-Dichlorobenzene	40.000	40.032	-0.1	112	0.00
13 C	1,4-Dichlorobenzene	40.000	39.484	1.3	110	0.00
14	1,2-Dichlorobenzene	40.000	38.878	2.8	111	0.00
15	Benzyl Alcohol	40.000	39.792	0.5	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.978	7.6	103	0.00
17	2-Methylphenol	40.000	39.303	1.7	109	0.00
18	Hexachloroethane	40.000	40.303	-0.8	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.564	6.1	109	0.00
20	3+4-Methylphenols	40.000	37.757	5.6	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	39.461	1.3	110	0.00
23 S	Nitrobenzene-d5	80.000	81.358	-1.7	109	0.00
24	Nitrobenzene	40.000	40.140	-0.4	107	0.00
25	Isophorone	40.000	39.751	0.6	109	0.00
26 C	2-Nitrophenol	40.000	41.029	-2.6	107	0.00
27	2,4-Dimethylphenol	40.000	39.458	1.4	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.159	2.1	105	0.00
29 C	2,4-Dichlorophenol	40.000	40.010	-0.0	107	0.00
30	1,2,4-Trichlorobenzene	40.000	40.652	-1.6	110	0.00
31	Naphthalene	40.000	39.445	1.4	107	0.00
32	Benzoic acid	40.000	35.784	10.5	89	0.00
33	4-Chloroaniline	40.000	39.594	1.0	106	0.00
34 C	Hexachlorobutadiene	40.000	41.805	-4.5	113	0.00
35	Caprolactam	40.000	37.533	6.2	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.366	-0.9	109	0.00
37	2-Methylnaphthalene	40.000	38.891	2.8	107	0.00
38	1-Methylnaphthalene	40.000	38.891	2.8	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.927	-4.8	111	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.712	5.7	94	0.00
42 S	2,4,6-Tribromophenol	80.000	83.264	-4.1	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.933	-2.3	106	0.00
44	2,4,5-Trichlorophenol	40.000	40.258	-0.6	105	0.00
45 S	2-Fluorobiphenyl	80.000	80.680	-0.9	109	0.00
46	1,1'-Biphenyl	40.000	40.296	-0.7	107	0.00
47	2-Chloronaphthalene	40.000	39.864	0.3	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063023\
 Data File : BF134051.D
 Acq On : 30 Jun 2023 14:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 30 18:36:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.069	-2.7	106	0.00
49	Acenaphthylene	40.000	39.381	1.5	104	0.00
50	Dimethylphthalate	40.000	39.896	0.3	106	0.00
51	2,6-Dinitrotoluene	40.000	40.975	-2.4	105	0.00
52 C	Acenaphthene	40.000	39.588	1.0	105	0.00
53	3-Nitroaniline	40.000	40.013	-0.0	104	0.00
54 P	2,4-Dinitrophenol	40.000	40.673	-1.7	108	0.00
55	Dibenzofuran	40.000	39.334	1.7	103	0.00
56 P	4-Nitrophenol	40.000	38.433	3.9	98	0.00
57	2,4-Dinitrotoluene	40.000	40.535	-1.3	104	0.00
58	Fluorene	40.000	39.848	0.4	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.678	0.8	105	0.00
60	Diethylphthalate	40.000	40.651	-1.6	107	0.00
61	4-Chlorophenyl-phenylether	40.000	40.185	-0.5	107	0.00
62	4-Nitroaniline	40.000	39.508	1.2	103	0.00
63	Azobenzene	40.000	40.450	-1.1	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.008	-7.5	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.769	3.1	103	0.00
67	4-Bromophenyl-phenylether	40.000	40.591	-1.5	106	0.00
68	Hexachlorobenzene	40.000	40.835	-2.1	109	0.00
69	Atrazine	40.000	37.774	5.6	105	0.00
70 C	Pentachlorophenol	40.000	40.359	-0.9	102	0.00
71	Phenanthrene	40.000	39.303	1.7	106	0.00
72	Anthracene	40.000	38.871	2.8	105	0.00
73	Carbazole	40.000	39.262	1.8	104	0.00
74	Di-n-butylphthalate	40.000	41.118	-2.8	108	0.00
75 C	Fluoranthene	40.000	39.338	1.7	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	46.452	-16.1	117	0.00
78	Pyrene	40.000	40.891	-2.2	106	0.00
79 S	Terphenyl-d14	80.000	83.709	-4.6	112	0.00
80	Butylbenzylphthalate	40.000	43.450	-8.6	114	0.00
81	Benzo(a)anthracene	40.000	40.051	-0.1	110	0.00
82	3,3'-Dichlorobenzidine	40.000	40.009	-0.0	111	0.00
83	Chrysene	40.000	39.153	2.1	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.153	-7.9	115	0.00
85 c	Di-n-octyl phthalate	40.000	41.014	-2.5	110	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.533	-6.3	94	0.00
88	Benzo(b)fluoranthene	40.000	41.377	-3.4	99	0.00
89	Benzo(k)fluoranthene	40.000	41.090	-2.7	102	0.00
90 C	Benzo(a)pyrene	40.000	40.684	-1.7	98	0.00
91	Dibenzo(a,h)anthracene	40.000	42.726	-6.8	94	0.00
92	Benzo(g,h,i)perylene	40.000	41.894	-4.7	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063023\
Data File : BF134051.D
Acq On : 30 Jun 2023 14:12
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 30 18:36:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0