

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070218\
 Data File : BF107052.D
 Acq On : 2 Jul 2018 18:50
 Operator : JU/SJ
 Sample : J3629-15
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 A0C22-01

Manual Integrations
 APPROVED

Sohil
 7/3/2018 5:30:03 PM

Quant Time: Jul 03 09:11:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	63533	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	225295	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	97121	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	190307	20.00	ng	0.00
75) Chrysene-d12	14.16	240	146978	20.00	ng	0.00
86) Perylene-d12	15.71	264	105447	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	385878	102.85	ng	0.00
7) Phenol-d6	6.61	99	462113	98.63	ng	0.00
23) Nitrobenzene-d5	7.52	82	255863	63.11	ng	-0.01
41) 2,4,6-Tribromophenol	10.80	330	124399	110.51	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	419361	70.58	ng	-0.01
78) Terphenyl-d14	13.08	244	481451	79.35	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.71	163	57696	7.921	ng	Qvalue # 98
70) Phenanthrene	11.52	178	53088	5.784	ng	97
74) Fluoranthene	12.72	202	105579	10.436	ng	95
77) Pyrene	12.95	202	117449	12.118	ng	97
80) Benzo(a)anthracene	14.15	228	40429	4.513	ng	94
82) Chrysene	14.18	228	40254	4.884	ng	97
83) Bis(2-ethylhexyl)phthalate	14.10	149	36891	7.241	ng	# 95
85) Indeno(1,2,3-cd)pyrene	17.31	276	17361	2.482	ng	# 88
87) Benzo(b)fluoranthene	15.25	252	34627m	5.286	ng	
88) Benzo(k)fluoranthene	15.28	252	12648m	2.028	ng	
89) Benzo(a)pyrene	15.64	252	28657	4.921	ng	# 91
91) Benzo(a,h,i)perylene	17.80	276	21817	4.523	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070218\
Data File : BF107052.D
Acq On : 2 Jul 2018 18:50
Operator : JU/SJ
Sample : J3629-15
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
A0C22-01

Manual Integrations
APPROVED

Sohil
7/3/2018 5:30:03 PM

Quant Time: Jul 03 09:11:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration

