


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070218\
 Data File : BF107057.D
 Acq On : 2 Jul 2018 21:34
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF061918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 29 14:38:29 2018



AutoFind: Scans 1595, 1596, 1597; Background Corrected with Scan 1590

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.4	18029	PASS
68	69	0.00	2	1.4	239	PASS
69	198	0.00	100	26.4	17359	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	36.5	24003	PASS
197	198	0.00	2	0.3	219	PASS
198	198	100	100	100.0	65717	PASS
199	198	5	9	6.7	4435	PASS
275	198	10	60	27.8	18270	PASS
365	198	1	100	3.3	2181	PASS
441	443	0.01	100	81.7	10019	PASS
442	198	50	100	99.2	65181	PASS
443	442	15	24	18.8	12266	PASS