

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF070224\  
 Data File : BF138430.D  
 Acq On : 02 Jul 2024 16:26  
 Operator : CG\JU  
 Sample : PB161821TB  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB161821TB

Quant Time: Jul 02 17:06:35 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF062824.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Jun 29 02:44:30 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.875  | 152  | 187333   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.151  | 136  | 767571   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.904  | 164  | 432252   | 20.000  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 11.392 | 188  | 683709   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 14.027 | 240  | 282888   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.492 | 264  | 221007   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.516  | 112  | 1549228  | 137.111 | ng    | 0.02     |
| 7) Phenol-d6                | 6.510  | 99   | 2073530  | 136.229 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.439  | 82   | 1371687  | 92.833  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.698 | 330  | 442093   | 125.184 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.233  | 172  | 2262814  | 83.312  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.980 | 244  | 1640481  | 79.471  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF070224\  
Data File : BF138430.D  
Acq On : 02 Jul 2024 16:26  
Operator : CG\JU  
Sample : PB161821TB  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB161821TB

Quant Time: Jul 02 17:06:35 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF062824.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Jun 29 02:44:30 2024  
Response via : Initial Calibration

