

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070224\
 Data File : BF138418.D
 Acq On : 02 Jul 2024 09:43
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 02 11:36:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 02:44:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	43.643	-9.1	113	-0.01
3	Pyridine	40.000	44.254	-10.6	112	0.00
4	n-Nitrosodimethylamine	40.000	42.542	-6.4	108	-0.01
5 S	2-Fluorophenol	80.000	82.717	-3.4	109	0.00
6	Aniline	40.000	42.024	-5.1	111	0.00
7 S	Phenol-d6	80.000	82.156	-2.7	109	0.00
8	2-Chlorophenol	40.000	41.297	-3.2	109	0.00
9	Benzaldehyde	40.000	40.491	-1.2	110	0.00
10 C	Phenol	40.000	41.659	-4.1	111	0.00
11	bis(2-Chloroethyl)ether	40.000	41.100	-2.8	108	0.00
12	1,3-Dichlorobenzene	40.000	39.667	0.8	104	0.00
13 C	1,4-Dichlorobenzene	40.000	39.815	0.5	106	0.00
14	1,2-Dichlorobenzene	40.000	39.538	1.2	105	0.00
15	Benzyl Alcohol	40.000	41.124	-2.8	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.675	-1.7	107	0.00
17	2-Methylphenol	40.000	41.180	-2.9	108	0.00
18	Hexachloroethane	40.000	40.527	-1.3	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.814	5.5	101	0.00
20	3+4-Methylphenols	40.000	41.375	-3.4	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	38.523	3.7	103	0.00
23 S	Nitrobenzene-d5	80.000	80.750	-0.9	105	0.00
24	Nitrobenzene	40.000	40.847	-2.1	106	0.00
25	Isophorone	40.000	39.155	2.1	104	0.00
26 C	2-Nitrophenol	40.000	43.157	-7.9	107	0.00
27	2,4-Dimethylphenol	40.000	41.122	-2.8	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.484	1.3	104	0.00
29 C	2,4-Dichlorophenol	40.000	39.912	0.2	103	0.00
30	1,2,4-Trichlorobenzene	40.000	38.065	4.8	100	0.00
31	Naphthalene	40.000	39.418	1.5	104	0.00
32	Benzoic acid	40.000	42.509	-6.3	114	-0.01
33	4-Chloroaniline	40.000	40.258	-0.6	107	0.00
34 C	Hexachlorobutadiene	40.000	37.521	6.2	98	0.00
35	Caprolactam	40.000	40.968	-2.4	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.799	0.5	105	0.00
37	2-Methylnaphthalene	40.000	38.318	4.2	102	0.00
38	1-Methylnaphthalene	40.000	38.628	3.4	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.135	4.7	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.766	8.1	89	0.00
42 S	2,4,6-Tribromophenol	80.000	77.801	2.7	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.689	3.3	98	0.00
44	2,4,5-Trichlorophenol	40.000	39.270	1.8	100	0.00
45 S	2-Fluorobiphenyl	80.000	73.107	8.6	98	0.00
46	1,1'-Biphenyl	40.000	38.133	4.7	100	0.00
47	2-Chloronaphthalene	40.000	38.837	2.9	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070224\
 Data File : BF138418.D
 Acq On : 02 Jul 2024 09:43
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 02 11:36:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 02:44:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.281	-5.7	108	0.00
49	Acenaphthylene	40.000	39.053	2.4	103	0.00
50	Dimethylphthalate	40.000	38.561	3.6	103	0.00
51	2,6-Dinitrotoluene	40.000	40.714	-1.8	105	0.00
52 C	Acenaphthene	40.000	39.634	0.9	104	0.00
53	3-Nitroaniline	40.000	41.233	-3.1	107	0.00
54 P	2,4-Dinitrophenol	40.000	42.254	-5.6	120	0.00
55	Dibenzofuran	40.000	39.036	2.4	103	0.00
56 P	4-Nitrophenol	40.000	42.895	-7.2	106	0.00
57	2,4-Dinitrotoluene	40.000	41.969	-4.9	105	0.00
58	Fluorene	40.000	36.891	7.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.164	-0.4	101	0.00
60	Diethylphthalate	40.000	38.392	4.0	102	0.00
61	4-Chlorophenyl-phenylether	40.000	35.928	10.2	100	0.00
62	4-Nitroaniline	40.000	40.406	-1.0	102	0.00
63	Azobenzene	40.000	40.510	-1.3	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.266	-5.7	114	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.786	3.0	103	0.00
67	4-Bromophenyl-phenylether	40.000	38.796	3.0	102	0.00
68	Hexachlorobenzene	40.000	38.022	4.9	101	0.00
69	Atrazine	40.000	36.750	8.1	90	0.00
70 C	Pentachlorophenol	40.000	40.531	-1.3	97	0.00
71	Phenanthrene	40.000	39.541	1.1	104	0.00
72	Anthracene	40.000	39.349	1.6	103	0.00
73	Carbazole	40.000	38.671	3.3	100	0.00
74	Di-n-butylphthalate	40.000	38.099	4.8	98	0.00
75 C	Fluoranthene	40.000	37.381	6.5	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	112	0.00
77	Benzydine	40.000	18.203	54.5#	252	0.00
78	Pyrene	40.000	34.093	14.8	95	0.00
79 S	Terphenyl-d14	80.000	66.037	17.5	90	0.00
80	Butylbenzylphthalate	40.000	35.753	10.6	95	0.00
81	Benzo(a)anthracene	40.000	39.413	1.5	111	0.00
82	3,3'-Dichlorobenzidine	40.000	42.139	-5.3	126	0.00
83	Chrysene	40.000	39.374	1.6	113	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.570	-3.9	113	0.00
85 c	Di-n-octyl phthalate	40.000	47.679	-19.2	134	0.00
86 I	Perylene-d12	20.000	20.000	0.0	121	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.820	5.4	110	0.00
88	Benzo(b)fluoranthene	40.000	40.982	-2.5	130	0.00
89	Benzo(k)fluoranthene	40.000	38.512	3.7	116	0.00
90 C	Benzo(a)pyrene	40.000	39.639	0.9	122	0.00
91	Dibenzo(a,h)anthracene	40.000	37.029	7.4	106	0.00
92	Benzo(g,h,i)perylene	40.000	37.712	5.7	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070224\
Data File : BF138418.D
Acq On : 02 Jul 2024 09:43
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 02 11:36:32 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 29 02:44:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0