

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070318\
 Data File : BF107086.D
 Acq On : 3 Jul 2018 17:06
 Operator : JU/SJ
 Sample : J3755-01DL 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PS-BASELINEWATERDL

Quant Time: Jul 03 17:45:50 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	54085	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	188885	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	82902	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	162811	20.00	ng	0.00
75) Chrysene-d12	14.15	240	145850	20.00	ng	-0.01
86) Perylene-d12	15.69	264	101997	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	95075	29.77	ng	0.01
7) Phenol-d6	6.64	99	69827	17.51	ng	0.02
23) Nitrobenzene-d5	7.52	82	141289	41.57	ng	-0.02
41) 2,4,6-Tribromophenol	10.80	330	66256	68.96	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	274346	49.93	ng	-0.01
78) Terphenyl-d14	13.08	244	347917	57.78	ng	0.00
Target Compounds						
22) Acetophenone	7.36	105	19563	4.629	ng	Qvalue # 98
25) Isophorone	7.78	82	282705	47.202	ng	99
49) Dimethylphthalate	9.71	163	62305	10.020	ng	97
59) Diethylphthalate	10.40	149	58820	9.801	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070318\
Data File : BF107086.D
Acq On : 3 Jul 2018 17:06
Operator : JU/SJ
Sample : J3755-01DL 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PS-BASELINEWATERDL

Quant Time: Jul 03 17:45:50 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration

