

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070318\
 Data File : BF107090.D
 Acq On : 3 Jul 2018 18:57
 Operator : JU/SJ
 Sample : J3798-07
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DRUM-13

Quant Time: Jul 05 03:51:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	54855	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	188578	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	82038	20.00	ng	0.00
63) Phenanthrene-d10	11.49	188	161310	20.00	ng	0.00
75) Chrysene-d12	14.15	240	143352	20.00	ng	-0.02
86) Perylene-d12	15.69	264	110640	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	133753	41.29	ng	0.00
7) Phenol-d6	6.60	99	155415	38.42	ng	-0.01
23) Nitrobenzene-d5	7.52	82	209933	61.87	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	41821	43.98	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	399599	81.91	ng	-0.01
78) Terphenyl-d14	13.08	244	496247	83.85	ng	0.00
Target Compounds						
10) Phenol	6.62	94	41857	9.066	ng	98
17) 2-Methylphenol	7.22	107	9704	3.191	ng	# 85
27) 2,4-Dimethylphenol	7.91	122	36129	14.292	ng	# 94
49) Dimethylphthalate	9.70	163	117495	19.096	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070318\
Data File : BF107090.D
Acq On : 3 Jul 2018 18:57
Operator : JU/SJ
Sample : J3798-07
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DRUM-13

Quant Time: Jul 05 03:51:17 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.95 min Scan# 785
Delta R.T. -0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Instrument :
BNA_F
ClientSampleId :
DRUM-13

Tot Ion:152 Resp: 54855
Ion Ratio Lower Upper
152 100
150 155.8 124.6 186.8
115 62.2 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 41.288 ng
RT: 5.60 min Scan# 555
Delta R.T. -0.00 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Tgt Ion:112 Resp: 133753
Ion Ratio Lower Upper
112 100
64 59.8 47.9 71.9
63 29.4 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 38.417 ng

RT: 6.60 min Scan# 725

Delta R.T. -0.01 min

Lab File: BF107090.D

Acq: 3 Jul 2018 18:57

Instrument :

BNA_F

ClientSampleId :

DRUM-13


Tot Ion: 99 Resp: 155415

Ion Ratio Lower Upper




99 100

42 17.6 14.5 21.7

71 36.1 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#10

Phenol

Concen: 9.066 ng

RT: 6.62 min Scan# 728

Delta R.T. -0.01 min

Lab File: BF107090.D

Acq: 3 Jul 2018 18:57


Tgt Ion: 94 Resp: 41857

Ion Ratio Lower Upper



94 100


65 26.2 7.9 47.9

66 35.2 16.2 56.2



Abundance Ion 94.00 (93.70 to 94.70): BF1
Ion 65.00 (64.70 to 65.70): BF1
Ion 66.00 (65.70 to 66.70): BF1









#17
2-Methylphenol
Concen: 3.191 ng
RT: 7.22 min Scan# 830
Delta R.T. -0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57




Instrument :
BNA_F
ClientSampleId :
DRUM-13


Tot Ion:107 Resp: 9704
Ion Ratio Lower Upper
107 100
108 102.6 91.7 137.5
77 57.9 33.8 50.8#
79 52.3 33.8 50.8#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.24 min Scan# 1003
Delta R.T. -0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Tgt Ion:136 Resp: 188578
Ion Ratio Lower Upper
136 100
137 10.9 8.9 13.3
54 8.2 6.6 9.8
68 6.6 5.0 7.4









#23
Nitrobenzene-d5
Concen: 61.867 ng
RT: 7.52 min Scan# 881
Delta R.T. -0.02 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Instrument :
BNA_F
ClientSampleId :
DRUM-13

Tot Ion: 82 Resp: 209933
Ion Ratio Lower Upper
82 100
128 43.7 36.1 54.1
54 49.4 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1






#27
2,4-Dimethylphenol
Concen: 14.292 ng
RT: 7.91 min Scan# 947
Delta R.T. 0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Tgt Ion:122 Resp: 36129
Ion Ratio Lower Upper
122 100
107 114.5 91.2 136.8
121 45.6 47.2 70.8#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E









#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.00 min Scan# 1303
Delta R.T. -0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57




Instrument :
BNA_F
ClientSampleId :
DRUM-13


Tot Ion:164 Resp: 82038
Ion Ratio Lower Upper
164 100
162 99.1 86.1 129.1
160 45.9 38.3 57.5



#41
2,4,6-Tribromophenol
Concen: 43.983 ng
RT: 10.79 min Scan# 1438
Delta R.T. -0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Tgt Ion:330 Resp: 41821
Ion Ratio Lower Upper
330 100
332 98.4 77.0 115.6
141 24.2 29.9 44.9#









#44
2-Fluorobiphenyl
Concen: 81.910 ng
RT: 9.31 min Scan# 1186
Delta R.T. -0.01 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57




Instrument :
BNA_F
ClientSampleId :
DRUM-13


Tot Ion:172 Resp: 399599
Ion Ratio Lower Upper
172 100
171 35.4 29.7 44.5
170 23.8 19.6 29.4



#49
Dimethylphthalate
Concen: 19.096 ng
RT: 9.70 min Scan# 1252
Delta R.T. -0.02 min
Lab File: BF107090.D
Acq: 3 Jul 2018 18:57

Tgt Ion:163 Resp: 117495
Ion Ratio Lower Upper
163 100
194 4.2 2.7 4.1#
164 11.2 8.2 12.2





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.49 min Scan# 1557

Delta R.T. -0.01 min

Lab File: BF107090.D

Acq: 3 Jul 2018 18:57

Instrument :

BNA_F

ClientSampleId :

DRUM-13



Tot Ion:188 Resp: 161310

Ion Ratio Lower Upper

188 100

94 8.2 8.5 12.7#



80 9.2 9.2 13.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.15 min Scan# 2008

Delta R.T. -0.02 min

Lab File: BF107090.D

Acq: 3 Jul 2018 18:57



Tgt Ion:240 Resp: 143352

Ion Ratio Lower Upper

240 100

120 8.8 9.5 14.3#


236 25.5 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
 Terphenyl-d14
 Concen: 83.854 ng
 RT: 13.08 min Scan# 1826
 Delta R.T. -0.01 min
 Lab File: BF107090.D
 Acq: 3 Jul 2018 18:57

Instrument :
 BNA_F
 ClientSampleId :
 DRUM-13

Tot Ion:244 Resp: 496247
 Ion Ratio Lower Upper
 244 100
 212 6.3 5.4 8.0
 122 8.5 11.0 16.4#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.69 min Scan# 2270
 Delta R.T. -0.03 min
 Lab File: BF107090.D
 Acq: 3 Jul 2018 18:57

Tgt Ion:264 Resp: 110640
 Ion Ratio Lower Upper
 264 100
 260 24.0 19.2 28.8
 265 21.8 17.5 26.3

