

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/3/2024	BNA_F	<u>BF138437.D</u>
Compound Name	Response	Retention Time
DDT	534449	13.551
DDD	44914	13.245
DDE	2075	12.91
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
46989	581438	8.08