

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070523\
 Data File : BF134073.D
 Acq On : 05 Jul 2023 18:31
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 06 00:20:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.493	0.494	-0.2	88	0.00
3	Pyridine	1.284	1.326	-3.3	91	-0.01
4	n-Nitrosodimethylamine	0.733	0.752	-2.6	89	-0.02
5 S	2-Fluorophenol	1.196	1.197	-0.1	90	0.00
6	Aniline	1.789	1.800	-0.6	88	0.00
7 S	Phenol-d6	1.470	1.435	2.4	89	0.00
8	2-Chlorophenol	1.214	1.195	1.6	89	0.00
9	Benzaldehyde	0.886	0.766	13.5	83	0.00
10 C	Phenol	1.546	1.538	0.5	90	0.00
11	bis(2-Chloroethyl)ether	1.138	1.103	3.1	86	0.00
12	1,3-Dichlorobenzene	1.294	1.271	1.8	88	0.00
13 C	1,4-Dichlorobenzene	1.307	1.286	1.6	87	0.00
14	1,2-Dichlorobenzene	1.224	1.218	0.5	91	0.00
15	Benzyl Alcohol	1.134	1.135	-0.1	87	0.00
16	2,2'-oxybis(1-Chloropropane	2.014	1.876	6.9	83	0.00
17	2-Methylphenol	1.087	1.061	2.4	86	0.00
18	Hexachloroethane	0.485	0.487	-0.4	89	0.00
19 P	n-Nitroso-di-n-propylamine	0.932	0.873	6.3	86	0.00
20	3+4-Methylphenols	1.319	1.286	2.5	88	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
22	Acetophenone	0.455	0.450	1.1	90	0.00
23 S	Nitrobenzene-d5	0.371	0.375	-1.1	89	0.00
24	Nitrobenzene	0.375	0.375	0.0	87	0.00
25	Isophorone	0.674	0.650	3.6	86	0.00
26 C	2-Nitrophenol	0.180	0.183	-1.7	87	0.00
27	2,4-Dimethylphenol	0.285	0.278	2.5	85	0.00
28	bis(2-Chloroethoxy)methane	0.390	0.372	4.6	83	0.00
29 C	2,4-Dichlorophenol	0.282	0.279	1.1	86	0.00
30	1,2,4-Trichlorobenzene	0.291	0.288	1.0	87	0.00
31	Naphthalene	0.966	0.935	3.2	86	0.00
32	Benzoic acid	0.213	0.200	6.1	76	-0.02
33	4-Chloroaniline	0.413	0.408	1.2	86	0.00
34 C	Hexachlorobutadiene	0.184	0.184	0.0	88	0.00
35	Caprolactam	0.093	0.089	4.3	83	-0.01
36 C	4-Chloro-3-methylphenol	0.317	0.318	-0.3	88	0.00
37	2-Methylnaphthalene	0.683	0.657	3.8	86	0.00
38	1-Methylnaphthalene	0.635	0.610	3.9	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	0.593	0.592	0.2	87	0.00
41 P	Hexachlorocyclopentadiene	0.281	0.239	14.9	70	0.00
42 S	2,4,6-Tribromophenol	0.251	0.251	0.0	87	0.00
43 C	2,4,6-Trichlorophenol	0.403	0.395	2.0	84	0.00
44	2,4,5-Trichlorophenol	0.474	0.457	3.6	83	0.00
45 S	2-Fluorobiphenyl	1.376	1.331	3.3	86	0.00
46	1,1'-Biphenyl	1.604	1.567	2.3	86	0.00
47	2-Chloronaphthalene	1.174	1.133	3.5	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070523\
 Data File : BF134073.D
 Acq On : 05 Jul 2023 18:31
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 06 00:20:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.428	0.432	-0.9	86	0.00
49	Acenaphthylene	2.005	1.928	3.8	84	0.00
50	Dimethylphthalate	1.452	1.399	3.7	84	0.00
51	2,6-Dinitrotoluene	0.313	0.307	1.9	83	0.00
52 C	Acenaphthene	1.164	1.128	3.1	85	0.00
53	3-Nitroaniline	0.375	0.371	1.1	85	0.00
54 P	2,4-Dinitrophenol	0.162	0.154	4.9	79	0.00
55	Dibenzofuran	1.818	1.748	3.9	83	0.00
56 P	4-Nitrophenol	0.262	0.244	6.9	78	0.00
57	2,4-Dinitrotoluene	0.413	0.410	0.7	85	0.00
58	Fluorene	1.415	1.372	3.0	86	0.00
59	2,3,4,6-Tetrachlorophenol	0.374	0.361	3.5	84	0.00
60	Diethylphthalate	1.513	1.454	3.9	84	0.00
61	4-Chlorophenyl-phenylether	0.682	0.660	3.2	85	0.00
62	4-Nitroaniline	0.378	0.371	1.9	85	0.00
63	Azobenzene	1.431	1.384	3.3	85	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	0.109	0.117	-7.3	84	-0.01
66 c	n-Nitrosodiphenylamine	0.639	0.638	0.2	86	0.00
67	4-Bromophenyl-phenylether	0.213	0.214	-0.5	85	0.00
68	Hexachlorobenzene	0.226	0.229	-1.3	87	0.00
69	Atrazine	0.177	0.174	1.7	88	0.00
70 C	Pentachlorophenol	0.135	0.134	0.7	80	0.00
71	Phenanthrene	1.007	0.987	2.0	85	0.00
72	Anthracene	1.047	1.040	0.7	86	0.00
73	Carbazole	0.959	0.965	-0.6	86	0.00
74	Di-n-butylphthalate	1.140	1.135	0.4	84	0.00
75 C	Fluoranthene	1.088	1.094	-0.6	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
77	Benzydine	0.476	0.565	-18.7	104	0.00
78	Pyrene	1.861	1.813	2.6	87	0.00
79 S	Terphenyl-d14	1.243	1.229	1.1	91	0.00
80	Butylbenzylphthalate	0.698	0.690	1.1	90	0.00
81	Benzo(a)anthracene	1.387	1.356	2.2	92	0.00
82	3,3'-Dichlorobenzidine	0.439	0.419	4.6	92	0.00
83	Chrysene	1.343	1.290	3.9	91	0.00
84	Bis(2-ethylhexyl)phthalate	0.802	0.775	3.4	89	0.00
85 c	Di-n-octyl phthalate	1.179	1.038	12.0	81	0.00
86 I	Perylene-d12	1.000	1.000	0.0	85	0.00
87	Indeno(1,2,3-cd)pyrene	1.388	1.406	-1.3	78	-0.01
88	Benzo(b)fluoranthene	1.176	1.162	1.2	83	0.00
89	Benzo(k)fluoranthene	1.197	1.207	-0.8	88	0.00
90 C	Benzo(a)pyrene	1.137	1.115	1.9	83	0.00
91	Dibenzo(a,h)anthracene	1.134	1.128	0.5	77	-0.01
92	Benzo(g,h,i)perylene	1.169	1.174	-0.4	78	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070523\
Data File : BF134073.D
Acq On : 05 Jul 2023 18:31
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 06 00:20:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0