

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070718\
 Data File : BF107216.D
 Acq On : 7 Jul 2018 19:32
 Operator : JU/SJ
 Sample : J3880-09
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SURCHARGE-SP-15

Manual Integrations
 APPROVED

Sohil
 7/9/2018 2:06:52 PM

Quant Time: Jul 09 05:03:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	47805	20.00	ng	-0.01
21) Naphthalene-d8	8.23	136	160051	20.00	ng	-0.02
38) Acenaphthene-d10	9.99	164	72188	20.00	ng	-0.02
63) Phenanthrene-d10	11.49	188	149930	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	116299	20.00	ng	-0.02
86) Perylene-d12	15.68	264	85154	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	307160	108.80	ng	0.00
7) Phenol-d6	6.60	99	364062	103.26	ng	-0.01
23) Nitrobenzene-d5	7.51	82	207168	71.93	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	97708	116.78	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	376392	88.94	ng	-0.02
78) Terphenyl-d14	13.07	244	475932	99.13	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	9.70	163	88400	16.327	ng	98
70) Phenanthrene	11.51	178	74197	10.260	ng	98
71) Anthracene	11.57	178	29284	3.967	ng	99
74) Fluoranthene	12.71	202	167658	21.035	ng	96
77) Pyrene	12.94	202	156413	20.395	ng	98
80) Benzo(a)anthracene	14.13	228	76031	10.727	ng	99
82) Chrysene	14.17	228	72370	11.098	ng	97
85) Indeno(1,2,3-cd)pyrene	17.28	276	36557	6.606	ng	# 87
87) Benzo(b)fluoranthene	15.22	252	80193m	15.160	ng	
88) Benzo(k)fluoranthene	15.25	252	28858m	5.729	ng	
89) Benzo(a)pyrene	15.62	252	55046	11.706	ng	96
90) Dibenzo(a,h)anthracene	17.28	278	11062	2.776	ng	# 81
91) Benzo(a,h,i)perylene	17.77	276	35443	9.099	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070718\
Data File : BF107216.D
Acq On : 7 Jul 2018 19:32
Operator : JU/SJ
Sample : J3880-09
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SURCHARGE-SP-15

Manual Integrations
APPROVED

Sohil
7/9/2018 2:06:52 PM

Quant Time: Jul 09 05:03:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration

