

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
 Data File : BF134148.D
 Acq On : 08 Jul 2023 14:55
 Operator : CG\JU
 Sample : 03523-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WALDWICK-COMP

Quant Time: Jul 09 00:37:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	159060	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	590041	20.000	ng	-0.01
39) Acenaphthene-d10	9.886	164	299951	20.000	ng	0.00
64) Phenanthrene-d10	11.380	188	639482	20.000	ng	0.00
76) Chrysene-d12	14.039	240	393138	20.000	ng	-0.01
86) Perylene-d12	15.545	264	339803	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	747538	78.615	ng	0.00
7) Phenol-d6	6.486	99	1004709	85.917	ng	0.00
23) Nitrobenzene-d5	7.410	82	672987	61.483	ng	-0.01
42) 2,4,6-Tribromophenol	10.680	330	359061	95.475	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	1214006	58.844	ng	0.00
79) Terphenyl-d14	12.980	244	1367131	55.967	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
Data File : BF134148.D
Acq On : 08 Jul 2023 14:55
Operator : CG\JU
Sample : 03523-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WALDWICK-COMP

Quant Time: Jul 09 00:37:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

