

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
 Data File : BF134168.D
 Acq On : 09 Jul 2023 03:24
 Operator : CG\JU
 Sample : 03509-12
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-WWS-03

Quant Time: Jul 09 04:19:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	119402	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	462195	20.000	ng	-0.01
39) Acenaphthene-d10	9.880	164	244909	20.000	ng	-0.01
64) Phenanthrene-d10	11.374	188	485659	20.000	ng	-0.01
76) Chrysene-d12	14.033	240	285226	20.000	ng	-0.02
86) Perylene-d12	15.539	264	237272	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	912016	127.769	ng	0.01
7) Phenol-d6	6.487	99	1050744	119.698	ng	0.00
23) Nitrobenzene-d5	7.410	82	838810	97.830	ng	-0.01
42) 2,4,6-Tribromophenol	10.680	330	478497	155.828	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	1564254	92.861	ng	0.00
79) Terphenyl-d14	12.980	244	2005122	113.141	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.592	163	47356	2.664	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
Data File : BF134168.D
Acq On : 09 Jul 2023 03:24
Operator : CG\JU
Sample : 03509-12
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-WWS-03

Quant Time: Jul 09 04:19:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

