

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
 Data File : BF134169.D
 Acq On : 09 Jul 2023 03:56
 Operator : CG\JU
 Sample : 03509-17
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-WWS-04

Quant Time: Jul 09 05:47:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	128033	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	513276	20.000	ng	-0.01
39) Acenaphthene-d10	9.881	164	266525	20.000	ng	-0.01
64) Phenanthrene-d10	11.375	188	517077	20.000	ng	-0.01
76) Chrysene-d12	14.033	240	305940	20.000	ng	-0.02
86) Perylene-d12	15.539	264	259391	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	953828	124.619	ng	0.00
7) Phenol-d6	6.487	99	1093257	116.145	ng	0.00
23) Nitrobenzene-d5	7.410	82	860812	90.404	ng	-0.01
42) 2,4,6-Tribromophenol	10.680	330	450059	134.680	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	1576546	86.001	ng	0.00
79) Terphenyl-d14	12.980	244	1899290	99.913	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
Data File : BF134169.D
Acq On : 09 Jul 2023 03:56
Operator : CG\JU
Sample : 03509-17
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-WWS-04

Quant Time: Jul 09 05:47:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

