

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
 Data File : BF134177.D
 Acq On : 09 Jul 2023 08:16
 Operator : CG\JU
 Sample : 03510-17
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-WWS-11

Quant Time: Jul 10 01:50:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	162620	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	626108	20.000	ng	-0.01
39) Acenaphthene-d10	9.880	164	305406	20.000	ng	-0.01
64) Phenanthrene-d10	11.375	188	648938	20.000	ng	-0.01
76) Chrysene-d12	14.033	240	347655	20.000	ng	-0.02
86) Perylene-d12	15.539	264	305880	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	1097195	112.861	ng	0.00
7) Phenol-d6	6.487	99	1284379	107.429	ng	0.00
23) Nitrobenzene-d5	7.410	82	1011493	87.085	ng	-0.01
42) 2,4,6-Tribromophenol	10.680	330	536130	140.012	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	1852556	88.191	ng	0.00
79) Terphenyl-d14	12.980	244	2218772	102.715	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
Data File : BF134177.D
Acq On : 09 Jul 2023 08:16
Operator : CG\JU
Sample : 03510-17
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-WWS-11

Quant Time: Jul 10 01:50:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

