

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
 Data File : BF134143.D
 Acq On : 08 Jul 2023 12:03
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 09 00:28:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	86	0.00
2	1,4-Dioxane	40.000	38.761	3.1	83	-0.02
3	Pyridine	40.000	38.841	2.9	84	-0.02
4	n-Nitrosodimethylamine	40.000	39.431	1.4	83	-0.03
5 S	2-Fluorophenol	80.000	77.782	2.8	85	0.00
6	Aniline	40.000	39.284	1.8	84	0.00
7 S	Phenol-d6	80.000	77.501	3.1	86	0.00
8	2-Chlorophenol	40.000	38.701	3.2	85	0.00
9	Benzaldehyde	40.000	39.175	2.1	84	0.00
10 C	Phenol	40.000	38.607	3.5	85	0.00
11	bis(2-Chloroethyl)ether	40.000	39.331	1.7	86	0.00
12	1,3-Dichlorobenzene	40.000	39.569	1.1	87	0.00
13 C	1,4-Dichlorobenzene	40.000	40.353	-0.9	87	0.00
14	1,2-Dichlorobenzene	40.000	39.009	2.5	87	0.00
15	Benzyl Alcohol	40.000	41.150	-2.9	88	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.950	5.1	83	0.00
17	2-Methylphenol	40.000	39.251	1.9	85	0.00
18	Hexachloroethane	40.000	41.581	-4.0	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.635	0.9	89	0.00
20	3+4-Methylphenols	40.000	39.150	2.1	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	39.607	1.0	89	0.00
23 S	Nitrobenzene-d5	80.000	81.902	-2.4	89	0.00
24	Nitrobenzene	40.000	40.452	-1.1	87	0.00
25	Isophorone	40.000	40.229	-0.6	89	0.00
26 C	2-Nitrophenol	40.000	40.675	-1.7	86	0.00
27	2,4-Dimethylphenol	40.000	40.069	-0.2	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.075	2.3	85	0.00
29 C	2,4-Dichlorophenol	40.000	40.283	-0.7	87	0.00
30	1,2,4-Trichlorobenzene	40.000	39.776	0.6	87	0.00
31	Naphthalene	40.000	39.038	2.4	86	0.00
32	Benzoic acid	40.000	40.790	-2.0	83	0.00
33	4-Chloroaniline	40.000	39.378	1.6	86	0.00
34 C	Hexachlorobutadiene	40.000	42.587	-6.5	93	0.00
35	Caprolactam	40.000	39.406	1.5	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.840	-2.1	90	0.00
37	2-Methylnaphthalene	40.000	39.435	1.4	88	0.00
38	1-Methylnaphthalene	40.000	39.470	1.3	89	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.110	-0.3	91	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.118	12.2	75	0.00
42 S	2,4,6-Tribromophenol	80.000	85.692	-7.1	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.004	2.5	87	0.00
44	2,4,5-Trichlorophenol	40.000	39.731	0.7	89	0.00
45 S	2-Fluorobiphenyl	80.000	79.207	1.0	92	0.00
46	1,1'-Biphenyl	40.000	39.240	1.9	89	0.00
47	2-Chloronaphthalene	40.000	38.747	3.1	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
 Data File : BF134143.D
 Acq On : 08 Jul 2023 12:03
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 09 00:28:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 18:35:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.177	-0.4	89	0.00
49	Acenaphthylene	40.000	38.602	3.5	88	0.00
50	Dimethylphthalate	40.000	39.935	0.2	91	0.00
51	2,6-Dinitrotoluene	40.000	40.262	-0.7	89	0.00
52 C	Acenaphthene	40.000	38.672	3.3	88	0.00
53	3-Nitroaniline	40.000	38.211	4.5	86	0.00
54 P	2,4-Dinitrophenol	40.000	39.529	1.2	90	0.00
55	Dibenzofuran	40.000	38.599	3.5	87	0.00
56 P	4-Nitrophenol	40.000	35.728	10.7	78	0.00
57	2,4-Dinitrotoluene	40.000	39.441	1.4	87	0.00
58	Fluorene	40.000	39.400	1.5	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.687	0.8	90	0.00
60	Diethylphthalate	40.000	40.665	-1.7	93	0.00
61	4-Chlorophenyl-phenylether	40.000	40.954	-2.4	94	0.00
62	4-Nitroaniline	40.000	37.776	5.6	85	0.00
63	Azobenzene	40.000	40.015	-0.0	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.900	-9.7	92	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.164	2.1	90	0.00
67	4-Bromophenyl-phenylether	40.000	40.808	-2.0	92	0.00
68	Hexachlorobenzene	40.000	40.718	-1.8	94	0.00
69	Atrazine	40.000	42.407	-6.0	101	0.00
70 C	Pentachlorophenol	40.000	38.958	2.6	84	0.00
71	Phenanthrene	40.000	38.870	2.8	90	0.00
72	Anthracene	40.000	39.038	2.4	91	0.00
73	Carbazole	40.000	38.463	3.8	88	0.00
74	Di-n-butylphthalate	40.000	42.809	-7.0	96	0.00
75 C	Fluoranthene	40.000	39.747	0.6	93	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	46.375	-15.9	106	0.00
78	Pyrene	40.000	38.990	2.5	91	0.00
79 S	Terphenyl-d14	80.000	83.064	-3.8	100	0.00
80	Butylbenzylphthalate	40.000	44.348	-10.9	106	0.00
81	Benzo(a)anthracene	40.000	38.371	4.1	95	0.00
82	3,3'-Dichlorobenzidine	40.000	37.513	6.2	95	0.00
83	Chrysene	40.000	37.530	6.2	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	46.852	-17.1	113	0.00
85 c	Di-n-octyl phthalate	40.000	44.023	-10.1	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.623	-4.1	80	0.00
88	Benzo(b)fluoranthene	40.000	39.573	1.1	83	0.00
89	Benzo(k)fluoranthene	40.000	41.099	-2.7	89	0.00
90 C	Benzo(a)pyrene	40.000	38.824	2.9	81	0.00
91	Dibenzo(a,h)anthracene	40.000	42.502	-6.3	82	0.00
92	Benzo(g,h,i)perylene	40.000	41.704	-4.3	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070823\
Data File : BF134143.D
Acq On : 08 Jul 2023 12:03
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 09 00:28:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 30 18:35:29 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0