

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070918\
 Data File : BF107233.D
 Acq On : 9 Jul 2018 17:14
 Operator : JU/SJ
 Sample : J3880-07RE
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SURCHARGE-SP-14RE

Manual Integrations
 APPROVED

Sohil
 7/10/2018 2:46:31 PM

Quant Time: Jul 09 18:29:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	36493	20.00	ng	-0.01
21) Naphthalene-d8	8.23	136	123921	20.00	ng	-0.02
38) Acenaphthene-d10	9.99	164	54361	20.00	ng	-0.02
63) Phenanthrene-d10	11.49	188	105226	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	98775	20.00	ng	-0.02
86) Perylene-d12	15.70	264	77363	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	223715	103.81	ng	0.00
7) Phenol-d6	6.60	99	267084	99.24	ng	-0.02
23) Nitrobenzene-d5	7.51	82	161940	72.62	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	68565	108.82	ng	-0.01
44) 2-Fluorobiphenyl	9.30	172	289139	91.08	ng	-0.02
78) Terphenyl-d14	13.07	244	336664	82.56	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.70	163	58395	14.323	ng	Qvalue 98
70) Phenanthrene	11.51	178	41399	8.157	ng	98
71) Anthracene	11.57	178	13720	2.648	ng	97
74) Fluoranthene	12.71	202	96914	17.325	ng	94
77) Pyrene	12.94	202	93881	14.413	ng	99
80) Benzo(a)anthracene	14.14	228	57640	9.575	ng	98
82) Chrysene	14.18	228	50859	9.183	ng	97
85) Indeno(1,2,3-cd)pyrene	17.29	276	22390	4.764	ng	# 88
87) Benzo(b)fluoranthene	15.24	252	58254m	12.122	ng	
88) Benzo(k)fluoranthene	15.27	252	21204m	4.633	ng	
89) Benzo(a)pyrene	15.63	252	37746	8.835	ng	94
91) Benzo(a,h,i)perylene	17.79	276	24335	6.876	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070918\
Data File : BF107233.D
Acq On : 9 Jul 2018 17:14
Operator : JU/SJ
Sample : J3880-07RE
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SURCHARGE-SP-14RE

Manual Integrations
APPROVED

Sohil
7/10/2018 2:46:31 PM

Quant Time: Jul 09 18:29:33 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration

