

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
 Data File : BF134210.D
 Acq On : 11 Jul 2023 12:54
 Operator : CG\JU
 Sample : PB153930TB
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB153930TB

Quant Time: Jul 12 01:03:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 11 13:06:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	123979	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	495677	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	199959	20.000	ng	0.00
64) Phenanthrene-d10	11.374	188	400966	20.000	ng	0.00
76) Chrysene-d12	14.039	240	215890	20.000	ng	0.00
86) Perylene-d12	15.545	264	208363	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	916658	123.679	ng	0.01
7) Phenol-d6	6.486	99	1075011	117.941	ng	0.00
23) Nitrobenzene-d5	7.410	82	726470	79.004	ng	0.00
42) 2,4,6-Tribromophenol	10.680	330	304578	121.487	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	1175463	85.467	ng	0.00
79) Terphenyl-d14	12.980	244	1278142	95.283	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
Data File : BF134210.D
Acq On : 11 Jul 2023 12:54
Operator : CG\JU
Sample : PB153930TB
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB153930TB

Quant Time: Jul 12 01:03:15 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 11 13:06:48 2023
Response via : Initial Calibration

