

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
 Data File : BF134224.D
 Acq On : 11 Jul 2023 21:23
 Operator : CG\JU
 Sample : 03509-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-WWS-02

Quant Time: Jul 12 01:56:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 11 13:06:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	107818	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	421959	20.000	ng	0.00
39) Acenaphthene-d10	9.881	164	217799	20.000	ng	0.00
64) Phenanthrene-d10	11.375	188	418162	20.000	ng	0.00
76) Chrysene-d12	14.033	240	219653	20.000	ng	0.00
86) Perylene-d12	15.539	264	209832	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	430081	66.726	ng	0.00
7) Phenol-d6	6.475	99	520181	65.624	ng	-0.01
23) Nitrobenzene-d5	7.404	82	351022	44.843	ng	0.00
42) 2,4,6-Tribromophenol	10.675	330	171071	62.646	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	598889	39.978	ng	0.00
79) Terphenyl-d14	12.969	244	622885	45.639	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
Data File : BF134224.D
Acq On : 11 Jul 2023 21:23
Operator : CG\JU
Sample : 03509-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-WWS-02

Quant Time: Jul 12 01:56:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 11 13:06:48 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.840 min Scan# 80
Delta R.T. -0.005 min
Lab File: BF134224.D
Acq: 11 Jul 2023 21:23




Instrument :
BNA_F
ClientSampleId :
WC-WWS-02


Tgt Ion:152 Resp: 107818
Ion Ratio Lower Upper
152 100
150 155.2 124.5 186.7
115 63.4 50.2 75.2



#5
2-Fluorophenol
Concen: 66.726 ng
RT: 5.475 min Scan# 576
Delta R.T. 0.000 min
Lab File: BF134224.D
Acq: 11 Jul 2023 21:23

Tgt Ion:112 Resp: 430081
Ion Ratio Lower Upper
112 100
64 62.1 52.1 78.1
63 35.9 28.6 42.8









#7
Phenol-d6
Concen: 65.624 ng
RT: 6.475 min Scan# 74
Delta R.T. -0.012 min
Lab File: BF134224.D
Acq: 11 Jul 2023 21:23




Instrument :
BNA_F
ClientSampleId :
WC-WWS-02


Tgt Ion: 99 Resp: 520181
Ion Ratio Lower Upper
99 100
42 24.9 19.8 29.6
71 42.3 31.9 47.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.122 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BF134224.D
Acq: 11 Jul 2023 21:23

Tgt Ion: 136 Resp: 421959
Ion Ratio Lower Upper
136 100
137 10.7 8.6 12.8
54 8.1 6.5 9.7
68 5.0 4.2 6.2





#23

Nitrobenzene-d5

Concen: 44.843 ng

RT: 7.404 min Scan# 904

Delta R.T. -0.006 min

Lab File: BF134224.D

Acq: 11 Jul 2023 21:23

Instrument :

BNA_F

ClientSampleId :

WC-WWS-02





Tgt Ion: 82 Resp: 351022

Ion Ratio Lower Upper

82 100

128 40.8 32.5 48.7

54 58.0 48.2 72.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.881 min Scan# 1325

Delta R.T. -0.006 min

Lab File: BF134224.D

Acq: 11 Jul 2023 21:23




Tgt Ion:164 Resp: 217799


Ion Ratio Lower Upper

164 100

162 101.6 80.0 120.0

160 45.4 35.9 53.9









#42
2,4,6-Tribromophenol
Concen: 62.646 ng
RT: 10.675 min Scan# 1460
Delta R.T. -0.006 min
Lab File: BF134224.D
Acq: 11 Jul 2023 21:23




Instrument :
BNA_F
ClientSampleId :
WC-WWS-02


Tgt Ion:330 Resp: 171071
Ion Ratio Lower Upper
330 100
332 95.9 77.0 115.6
141 32.7 26.3 39.5



#45
2-Fluorobiphenyl
Concen: 39.978 ng
RT: 9.198 min Scan# 1209
Delta R.T. -0.006 min
Lab File: BF134224.D
Acq: 11 Jul 2023 21:23

Tgt Ion:172 Resp: 598889
Ion Ratio Lower Upper
172 100
171 36.1 28.6 43.0
170 23.5 19.4 29.0





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.375 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF134224.D

Acq: 11 Jul 2023 21:23

Instrument :

BNA_F

ClientSampleId :

WC-WWS-02





Tgt Ion:188 Resp: 418162

Ion Ratio Lower Upper

188 100

94 8.5 6.4 9.6

80 8.9 7.3 10.9



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.033 min Scan# 2031

Delta R.T. -0.006 min

Lab File: BF134224.D

Acq: 11 Jul 2023 21:23




Tgt Ion:240 Resp: 219653


Ion Ratio Lower Upper

240 100

120 10.6 7.4 11.0

236 26.4 21.0 31.6









#79
 Terphenyl-d14
 Concen: 45.639 ng
 RT: 12.969 min Scan# 1850
 Delta R.T. -0.006 min
 Lab File: BF134224.D
 Acq: 11 Jul 2023 21:23

Instrument :
 BNA_F
 ClientSampleId :
 WC-WWS-02

Tgt Ion:244 Resp: 622885
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.0 9.0
 122 10.5 8.3 12.5



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.539 min Scan# 2287
 Delta R.T. -0.006 min
 Lab File: BF134224.D
 Acq: 11 Jul 2023 21:23

Tgt Ion:264 Resp: 209832
 Ion Ratio Lower Upper
 264 100
 260 21.7 19.5 29.3
 265 21.1 16.6 25.0

