



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
 Data File : BF134216.D
 Acq On : 11 Jul 2023 16:15
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 11 13:06:48 2023




AutoFind: Scans 1622, 1623, 1624; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.7	43314	PASS
68	69	0.00	2	1.9	726	PASS
69	198	0.00	100	36.0	38319	PASS
70	69	0.00	2	0.5	180	PASS
127	198	10	80	47.4	50507	PASS
197	198	0.00	2	0.4	421	PASS
198	198	100	100	100.0	106453	PASS
199	198	5	9	6.7	7184	PASS
275	198	10	60	31.5	33585	PASS
365	198	1	100	4.2	4453	PASS
441	198	0.01	100	18.4	19596	PASS
442	442	50	100	100.0	124459	PASS
443	442	15	24	19.3	23960	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
Data File : BF134216.D
Acq On : 11 Jul 2023 16:15
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jul 12 01:37:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 11 13:06:48 2023
Response via : Initial Calibration



(70) Pentachlorophenol (C)

11.180min (-0.006) 31765.85 ng




response 94312

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.80	71.30
264.00	63.30	64.41
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071123\
 Data File : BF134216.D
 Acq On : 11 Jul 2023 16:15
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jul 12 01:37:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 11 13:06:48 2023
 Response via : Initial Calibration



TIC: BF134216.D\data.ms

(77) Benzidine

12.727min (0.000) 0.00 ng

response 502066

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.32
183.00	11.00	10.69
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/11/2023	BNA_F	<u>BF134216.D</u>
Compound Name	Response	Retention Time
DDT	250138	13.551
DDD	5972	13.251
DDE	1125	12.904
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7097	257235	2.76

Instrument :
BNA_F
ClientSampleId :
DFTPP