

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071218\
 Data File : BF107353.D
 Acq On : 12 Jul 2018 21:33
 Operator : JU/SJ
 Sample : J3933-07
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP-7-N2

Manual Integrations
 APPROVED

Sohil
 7/13/2018 11:17:18 AM

Quant Time: Jul 13 04:02:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 11 16:39:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.93	152	54042	20.00	ng	0.00
21) Naphthalene-d8	8.21	136	183650	20.00	ng	0.00
38) Acenaphthene-d10	9.98	164	77545	20.00	ng	0.00
63) Phenanthrene-d10	11.47	188	150408	20.00	ng	0.00
75) Chrysene-d12	14.13	240	157394	20.00	ng	-0.01
86) Perylene-d12	15.67	264	111293	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	360239	112.88	ng	0.01
7) Phenol-d6	6.58	99	429782	107.84	ng	0.00
23) Nitrobenzene-d5	7.50	82	263595	79.77	ng	0.00
41) 2,4,6-Tribromophenol	10.78	330	100919	112.29	ng	0.00
44) 2-Fluorobiphenyl	9.29	172	440988	98.62	ng	0.00
78) Terphenyl-d14	13.06	244	510968	78.64	ng	0.00
Target Compounds						
10) Phenol	6.60	94	12121	2.665	ng	80
49) Dimethylphthalate	9.68	163	74432	12.798	ng	99
70) Phenanthrene	11.50	178	59769	8.239	ng	97
74) Fluoranthene	12.70	202	130859	16.366	ng	95
77) Pyrene	12.93	202	131466	12.666	ng	99
80) Benzo(a)anthracene	14.12	228	72794	7.588	ng	98
82) Chrysene	14.16	228	67899	7.694	ng	98
85) Indeno(1,2,3-cd)pyrene	17.26	276	28799m	3.845	ng	
87) Benzo(b)fluoranthene	15.21	252	71043m	10.276	ng	
88) Benzo(k)fluoranthene	15.24	252	21303m	3.236	ng	
89) Benzo(a)pyrene	15.60	252	47326	7.700	ng	97
91) Benzo(a,h,i)perylene	17.75	276	28843	5.666	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071218\
Data File : BF107353.D
Acq On : 12 Jul 2018 21:33
Operator : JU/SJ
Sample : J3933-07
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EP-7-N2

Quant Time: Jul 13 04:02:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 11 16:39:37 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
7/13/2018 11:17:18 AM

