

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071323\
 Data File : BF134266.D
 Acq On : 13 Jul 2023 17:29
 Operator : CG\JU
 Sample : 03569-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WALL

Quant Time: Jul 13 18:15:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 11 13:06:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	80146	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	310273	20.000	ng	# 0.00
39) Acenaphthene-d10	9.880	164	152302	20.000	ng	0.00
64) Phenanthrene-d10	11.374	188	294259	20.000	ng	0.00
76) Chrysene-d12	14.033	240	221767	20.000	ng	0.00
86) Perylene-d12	15.539	264	206389	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	477730	99.709	ng	0.00
7) Phenol-d6	6.481	99	571551	97.000	ng	0.00
23) Nitrobenzene-d5	7.404	82	384348	66.775	ng	0.00
42) 2,4,6-Tribromophenol	10.674	330	192356	100.733	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	737518	70.404	ng	0.00
79) Terphenyl-d14	12.968	244	882025	64.011	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071323\
Data File : BF134266.D
Acq On : 13 Jul 2023 17:29
Operator : CG\JU
Sample : 03569-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WALL

Quant Time: Jul 13 18:15:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 11 13:06:48 2023
Response via : Initial Calibration

