

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071323\
 Data File : BF134270.D
 Acq On : 13 Jul 2023 19:48
 Operator : CG\JU
 Sample : 03358-02RE
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-1-(8-10)RE

Quant Time: Jul 14 01:44:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 11 13:06:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	70231	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	264005	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	129484	20.000	ng	0.00
64) Phenanthrene-d10	11.374	188	262477	20.000	ng	0.00
76) Chrysene-d12	14.033	240	175057	20.000	ng	0.00
86) Perylene-d12	15.539	264	164439	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	655970	156.240	ng	0.00
7) Phenol-d6	6.486	99	829231	160.601	ng	0.00
23) Nitrobenzene-d5	7.404	82	555702	113.465	ng	0.00
42) 2,4,6-Tribromophenol	10.674	330	299236	184.319	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	1019079	114.426	ng	0.00
79) Terphenyl-d14	12.974	244	1271776	116.923	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071323\
Data File : BF134270.D
Acq On : 13 Jul 2023 19:48
Operator : CG\JU
Sample : 03358-02RE
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-1-(8-10)RE

Quant Time: Jul 14 01:44:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 11 13:06:48 2023
Response via : Initial Calibration

