

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071519\
 Data File : BF115576.D
 Acq On : 16 Jul 2019 00:06
 Operator : HP/JU
 Sample : K3620-05 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CL-01-071219-C

Quant Time: Jul 16 05:43:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 12 14:03:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	184162	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	672949	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	304669	20.00	ng	-0.01
64) Phenanthrene-d10	11.40	188	500683	20.00	ng	-0.01
76) Chrysene-d12	14.03	240	354505	20.00	ng	-0.02
87) Perylene-d12	15.51	264	360384	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	484217	43.01	ng	0.00
7) Phenol-d6	6.50	99	648680	45.41	ng	-0.01
23) Nitrobenzene-d5	7.43	82	386645	33.41	ng	-0.02
42) 2,4,6-Tribromophenol	10.70	330	140227	39.43	ng	-0.01
45) 2-Fluorobiphenyl	9.24	172	656706	37.73	ng	-0.01
79) Terphenyl-d14	12.98	244	510959	26.59	ng	-0.02
Target Compounds						
50) Dimethylphthalate	9.63	163	59578	2.599	ng	98
75) Fluoranthene	12.61	202	90156	3.324	ng	100
78) Pyrene	12.84	202	102752	3.421	ng	96
83) Chrysene	14.06	228	53091	2.264	ng	96
88) Benzo(b)fluoranthene	15.07	252	50654m	2.183	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071519\
Data File : BF115576.D
Acq On : 16 Jul 2019 00:06
Operator : HP/JU
Sample : K3620-05 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-01-071219-C

Quant Time: Jul 16 05:43:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 12 14:03:52 2019
Response via : Initial Calibration

