

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071524\
 Data File : BF138569.D
 Acq On : 15 Jul 2024 14:59
 Operator : CG\JU
 Sample : PB162012TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB162012TB

Quant Time: Jul 15 15:39:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 16:58:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	124599	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	502869	20.000	ng	0.00
39) Acenaphthene-d10	9.898	164	321659	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	450624	20.000	ng	0.00
76) Chrysene-d12	14.022	240	247407	20.000	ng	0.00
86) Perylene-d12	15.492	264	263328	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	958547	121.839	ng	0.01
7) Phenol-d6	6.504	99	1215475	116.137	ng	0.00
23) Nitrobenzene-d5	7.428	82	807402	78.831	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	310411	103.256	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	1735123	84.308	ng	0.00
79) Terphenyl-d14	12.975	244	1930054	127.090	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071524\
Data File : BF138569.D
Acq On : 15 Jul 2024 14:59
Operator : CG\JU
Sample : PB162012TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB162012TB

Quant Time: Jul 15 15:39:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 16:58:40 2024
Response via : Initial Calibration

