

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071818\
 Data File : BF107499.D
 Acq On : 19 Jul 2018 7:09
 Operator : JU/SJ
 Sample : J4027-07
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WALL-3-159

Quant Time: Jul 19 08:24:21 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 18 11:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.98	152	119333	20.00	ng	0.00
21) Naphthalene-d8	8.27	136	423266	20.00	ng	0.00
38) Acenaphthene-d10	10.02	164	215775	20.00	ng	0.00
63) Phenanthrene-d10	11.52	188	409091	20.00	ng	0.00
75) Chrysene-d12	14.17	240	323017	20.00	ng	-0.01
86) Perylene-d12	15.71	264	290661	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	901404	114.72	ng	0.02
7) Phenol-d6	6.64	99	1193873	117.52	ng	0.02
23) Nitrobenzene-d5	7.55	82	815093	102.69	ng	0.00
41) 2,4,6-Tribromophenol	10.82	330	241471	127.37	ng	0.00
44) 2-Fluorobiphenyl	9.35	172	1263694	99.50	ng	0.00
78) Terphenyl-d14	13.11	244	1240802	95.44	ng	0.00
Target Compounds						
25) Isophorone	7.80	82	68637	4.363	ng	Qvalue # 84
35) Caprolactam	8.68	113	40240	18.800	ng	92
49) Dimethylphthalate	9.73	163	269804	15.737	ng	99
59) Diethylphthalate	10.44	149	134036	8.047	ng	# 90
73) Di-n-butylphthalate	12.07	149	74411	3.368	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071818\
Data File : BF107499.D
Acq On : 19 Jul 2018 7:09
Operator : JU/SJ
Sample : J4027-07
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WALL-3-159

Quant Time: Jul 19 08:24:21 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 18 11:55:34 2018
Response via : Initial Calibration

