


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071818\
 Data File : BF107463.D
 Acq On : 18 Jul 2018 8:51
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071618.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 18 11:55:34 2018



AutoFind: Scans 1601, 1602, 1603; Background Corrected with Scan 1596

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.9	41562	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	27.7	41325	PASS
70	69	0.00	2	1.6	669	PASS
127	198	10	80	29.1	43465	PASS
197	198	0.00	2	0.8	1160	PASS
198	198	100	100	100.0	149218	PASS
199	198	5	9	6.9	10342	PASS
275	198	10	60	20.6	30808	PASS
365	198	1	100	2.1	3126	PASS
441	198	0.01	100	7.9	11847	PASS
442	198	50	100	54.5	81258	PASS
443	442	15	24	16.8	13641	PASS