

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071823\
 Data File : BF134350.D
 Acq On : 18 Jul 2023 21:56
 Operator : CG\JU
 Sample : 03608-02RE
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 73742RE

Quant Time: Jul 19 01:42:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 14 22:51:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	62120	20.000	ng	0.00
21) Naphthalene-d8	8.116	136	233800	20.000	ng	0.00
39) Acenaphthene-d10	9.886	164	89686	20.000	ng	0.00
64) Phenanthrene-d10	11.428	188	123757	20.000	ng	# 0.05
76) Chrysene-d12	14.057	240	152435	20.000	ng	0.02
86) Perylene-d12	15.545	264	161847	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	56776	15.289	ng	0.00
7) Phenol-d6	6.475	99	69421	15.201	ng	-0.01
23) Nitrobenzene-d5	7.398	82	44362	10.228	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	17296	15.381	ng	0.02
45) 2-Fluorobiphenyl	9.192	172	88116	14.284	ng	-0.01
79) Terphenyl-d14	13.004	244	79984	8.445	ng	0.03
Target Compounds						
71) Phenanthrene	11.451	178	157440	25.271	ng	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071823\
Data File : BF134350.D
Acq On : 18 Jul 2023 21:56
Operator : CG\JU
Sample : 03608-02RE
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
73742RE

Quant Time: Jul 19 01:42:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 14 22:51:27 2023
Response via : Initial Calibration

