



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071923\
 Data File : BF134354.D
 Acq On : 19 Jul 2023 12:29
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 14 22:51:27 2023






AutoFind: Scans 1621, 1622, 1623; Background Corrected with Scan 1615

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.4	42770	PASS
68	69	0.00	2	1.5	612	PASS
69	198	0.00	100	37.4	40611	PASS
70	69	0.00	2	0.5	195	PASS
127	198	10	80	46.8	50886	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	108677	PASS
199	198	5	9	6.9	7548	PASS
275	198	10	60	32.8	35640	PASS
365	198	1	100	4.6	5018	PASS
441	198	0.01	100	21.1	22929	PASS
442	442	50	100	100.0	146019	PASS
443	442	15	24	18.8	27480	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071923\
 Data File : BF134354.D
 Acq On : 19 Jul 2023 12:29
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jul 19 14:08:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 14 22:51:27 2023
 Response via : Initial Calibration



TIC: BF134354.D\data.ms

(70) Pentachlorophenol (C)

11.180min (0.000) 24915.29 ng




response 107765

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.80	71.93
264.00	63.30	63.77
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071923\
Data File : BF134354.D
Acq On : 19 Jul 2023 12:29
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jul 19 14:08:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 14 22:51:27 2023
Response via : Initial Calibration



TIC: BF134354.D\data.ms

(77) Benzidine

12.727min (0.000) 426374.76 ng

response 598573

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	13.69
183.00	11.00	11.08
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/19/2023	BNA_F	<u>BF134354.D</u>
Compound Name	Response	Retention Time
DDT	287822	13.545
DDD	7676	13.268
DDE	428	12.91
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8104	295926	2.74

Instrument :
BNA_F
ClientSampleId :
DFTPP