

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072018\
 Data File : BF107552.D
 Acq On : 21 Jul 2018 4:02
 Operator : JU/SJ
 Sample : J4043-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SD-BASELINE-WATER

Quant Time: Jul 21 05:38:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF072018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 20 16:29:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.98	152	262141	20.00	ng	0.00
21) Naphthalene-d8	8.26	136	909292	20.00	ng	0.00
38) Acenaphthene-d10	10.02	164	326587	20.00	ng	0.00
63) Phenanthrene-d10	11.51	188	680318	20.00	ng	0.00
75) Chrysene-d12	14.17	240	611982	20.00	ng	0.00
86) Perylene-d12	15.71	264	524383	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	165474	10.15	ng	0.00
7) Phenol-d6	6.61	99	167507	8.56	ng	0.00
23) Nitrobenzene-d5	7.54	82	170322	12.64	ng	-0.01
41) 2,4,6-Tribromophenol	10.81	330	45071	12.14	ng	0.00
44) 2-Fluorobiphenyl	9.33	172	273996	8.00	ng	0.00
78) Terphenyl-d14	13.09	244	316492	13.24	ng	0.00
Target Compounds						
10) Phenol	6.62	94	167818	7.289	ng	98
20) 3+4-Methylphenols	7.38	107	60546	3.434	ng	# 88
49) Dimethylphthalate	9.72	163	171148	6.953	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072018\
Data File : BF107552.D
Acq On : 21 Jul 2018 4:02
Operator : JU/SJ
Sample : J4043-03
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SD-BASELINE-WATER

Quant Time: Jul 21 05:38:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF072018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 20 16:29:27 2018
Response via : Initial Calibration

