


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072018\
 Data File : BF107518.D
 Acq On : 20 Jul 2018 12:25
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 20 16:29:27 2018



AutoFind: Scans 1599, 1600, 1601; Background Corrected with Scan 1593

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.1	119744	PASS
68	69	0.00	2	1.9	1761	PASS
69	198	0.00	100	34.1	92593	PASS
70	69	0.00	2	0.4	380	PASS
127	198	10	80	47.5	129168	PASS
197	198	0.00	2	0.7	1807	PASS
198	198	100	100	100.0	271829	PASS
199	198	5	9	6.5	17546	PASS
275	198	10	60	27.9	75856	PASS
365	198	1	100	3.5	9625	PASS
441	198	0.01	100	14.6	39658	PASS
442	198	50	100	90.6	246144	PASS
443	442	15	24	18.8	46240	PASS