



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072021\
 Data File : BF124626.D
 Acq On : 20 Jul 2021 11:09
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 20 13:36:28 2021






AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.4	4622	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.8	3161	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.6	4733	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	9353	PASS
199	198	5	9	5.8	544	PASS
275	198	10	60	24.0	2244	PASS
365	198	1	100	1.5	139	PASS
441	198	0.01	100	13.1	1224	PASS
442	442	50	100	100.0	7107	PASS
443	442	15	24	18.5	1318	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072021\
 Data File : BF124626.D
 Acq On : 20 Jul 2021 11:09
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jul 20 11:46:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 19 11:33:09 2021
 Response via : Initial Calibration



TIC: BF124626.D\data.ms

(77) Benzidine

12.763min (-0.012) 0.00 ng




response 19214

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	18.08#
183.00	10.10	11.52
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072021\
Data File : BF124626.D
Acq On : 20 Jul 2021 11:09
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jul 20 11:46:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 19 11:33:09 2021
Response via : Initial Calibration



TIC: BF124626.D\data.ms

(70) Pentachlorophenol (C)

11.227min (-0.018) 0.00 ng

response 8004

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	64.78
264.00	69.90	62.63
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/20/2021	BNA_F	<u>BF124626.D</u>
Compound Name	Response	Retention Time
DDT	18468	13.59
DDD	0	13.3
DDE	0	12.71
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
0	18468	0.00