

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072023\
 Data File : BF134386.D
 Acq On : 20 Jul 2023 17:39
 Operator : CG\JU
 Sample : 03646-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GATE-1

Quant Time: Jul 20 18:10:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 14 22:51:27 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/21/2023
 Supervised By :mohammad ahmed 07/21/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.834	152	64114	20.000	ng	-0.01
21) Naphthalene-d8	8.116	136	256179	20.000	ng	# 0.00
39) Acenaphthene-d10	9.869	164	132647	20.000	ng	-0.01
64) Phenanthrene-d10	11.369	188	268937	20.000	ng	0.00
76) Chrysene-d12	14.027	240	152926	20.000	ng	-0.01
86) Perylene-d12	15.533	264	156150	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	358114	93.434	ng	0.01
7) Phenol-d6	6.481	99	397201	84.267	ng	0.00
23) Nitrobenzene-d5	7.398	82	317014	66.706	ng	-0.01
42) 2,4,6-Tribromophenol	10.669	330	178415	107.277	ng	-0.01
45) 2-Fluorobiphenyl	9.192	172	651940	71.457	ng	-0.01
79) Terphenyl-d14	12.968	244	821196	86.424	ng	0.00
Target Compounds						
32) Benzoic acid	7.863	122	10797m	3.951	ng	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072023\
Data File : BF134386.D
Acq On : 20 Jul 2023 17:39
Operator : CG\JU
Sample : 03646-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GATE-1

Quant Time: Jul 20 18:10:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 14 22:51:27 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/21/2023
Supervised By :mohammad ahmed 07/21/2023

