

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072123\
 Data File : BF134392.D
 Acq On : 21 Jul 2023 09:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 21 13:26:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 14 22:51:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	51	0.00
2	1,4-Dioxane	40.000	39.159	2.1	50	0.01
3	Pyridine	40.000	37.200	7.0	47	0.01
4	n-Nitrosodimethylamine	40.000	37.368	6.6	47	0.02
5 S	2-Fluorophenol	80.000	81.133	-1.4	53	0.00
6	Aniline	40.000	39.393	1.5	50	0.00
7 S	Phenol-d6	80.000	81.237	-1.5	53	0.00
8	2-Chlorophenol	40.000	40.441	-1.1	53	0.00
9	Benzaldehyde	40.000	42.210	-5.5	53	0.00
10 C	Phenol	40.000	39.436	1.4	51	0.00
11	bis(2-Chloroethyl)ether	40.000	38.207	4.5	49	0.00
12	1,3-Dichlorobenzene	40.000	40.626	-1.6	53	0.00
13 C	1,4-Dichlorobenzene	40.000	40.920	-2.3	53	0.00
14	1,2-Dichlorobenzene	40.000	40.840	-2.1	54	0.00
15	Benzyl Alcohol	40.000	41.977	-4.9	53	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.102	19.7	41	0.00
17	2-Methylphenol	40.000	41.016	-2.5	53	0.00
18	Hexachloroethane	40.000	41.721	-4.3	53	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.021	-2.6	55	0.00
20	3+4-Methylphenols	40.000	42.852	-7.1	56	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	52	0.00
22	Acetophenone	40.000	42.560	-6.4	57	0.00
23 S	Nitrobenzene-d5	80.000	84.264	-5.3	54	0.00
24	Nitrobenzene	40.000	40.735	-1.8	52	0.00
25	Isophorone	40.000	40.012	-0.0	53	0.00
26 C	2-Nitrophenol	40.000	42.467	-6.2	53	0.00
27	2,4-Dimethylphenol	40.000	39.215	2.0	51	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.742	3.1	50	0.00
29 C	2,4-Dichlorophenol	40.000	40.573	-1.4	52	0.00
30	1,2,4-Trichlorobenzene	40.000	41.705	-4.3	54	0.00
31	Naphthalene	40.000	40.769	-1.9	53	0.00
32	Benzoic acid	40.000	34.551	13.6	42	0.00
33	4-Chloroaniline	40.000	39.251	1.9	51	0.00
34 C	Hexachlorobutadiene	40.000	44.869	-12.2	58	0.00
35	Caprolactam	40.000	40.261	-0.7	51	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.219	-8.0	56	0.00
37	2-Methylnaphthalene	40.000	40.390	-1.0	54	0.00
38	1-Methylnaphthalene	40.000	40.968	-2.4	55	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	54	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.632	-4.1	57	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.173	-5.4	54	0.00
42 S	2,4,6-Tribromophenol	80.000	95.624	-19.5	65	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.784	3.0	52	0.00
44	2,4,5-Trichlorophenol	40.000	38.423	3.9	52	0.00
45 S	2-Fluorobiphenyl	80.000	84.821	-6.0	59	0.00
46	1,1'-Biphenyl	40.000	40.077	-0.2	55	0.00
47	2-Chloronaphthalene	40.000	39.712	0.7	54	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072123\
 Data File : BF134392.D
 Acq On : 21 Jul 2023 09:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 21 13:26:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 14 22:51:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.781	0.5	53	0.00
49	Acenaphthylene	40.000	38.863	2.8	53	0.00
50	Dimethylphthalate	40.000	40.705	-1.8	56	0.00
51	2,6-Dinitrotoluene	40.000	41.546	-3.9	55	0.00
52 C	Acenaphthene	40.000	40.779	-1.9	56	0.00
53	3-Nitroaniline	40.000	39.592	1.0	53	0.00
54 P	2,4-Dinitrophenol	40.000	41.135	-2.8	57	0.00
55	Dibenzofuran	40.000	40.349	-0.9	55	0.00
56 P	4-Nitrophenol	40.000	38.835	2.9	51	0.00
57	2,4-Dinitrotoluene	40.000	42.465	-6.2	57	0.00
58	Fluorene	40.000	42.089	-5.2	58	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.836	0.4	55	0.00
60	Diethylphthalate	40.000	41.250	-3.1	56	0.00
61	4-Chlorophenyl-phenylether	40.000	43.219	-8.0	60	0.00
62	4-Nitroaniline	40.000	35.388	11.5	48	0.00
63	Azobenzene	40.000	38.879	2.8	53	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	56	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.647	-6.6	55	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.329	1.7	56	0.00
67	4-Bromophenyl-phenylether	40.000	41.406	-3.5	58	0.00
68	Hexachlorobenzene	40.000	43.347	-8.4	62	0.00
69	Atrazine	40.000	40.414	-1.0	60	0.00
70 C	Pentachlorophenol	40.000	35.378	11.6	47	0.00
71	Phenanthrene	40.000	40.286	-0.7	58	0.00
72	Anthracene	40.000	39.943	0.1	58	0.00
73	Carbazole	40.000	40.669	-1.7	57	0.00
74	Di-n-butylphthalate	40.000	42.212	-5.5	59	0.00
75 C	Fluoranthene	40.000	43.008	-7.5	62	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	77	0.00
77	Benzydine	40.000	39.259	1.9	71	0.00
78	Pyrene	40.000	35.773	10.6	66	0.00
79 S	Terphenyl-d14	80.000	74.734	6.6	71	0.00
80	Butylbenzylphthalate	40.000	38.755	3.1	73	0.00
81	Benzo(a)anthracene	40.000	39.364	1.6	77	0.00
82	3,3'-Dichlorobenzidine	40.000	40.892	-2.2	81	0.00
83	Chrysene	40.000	38.720	3.2	76	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.136	-12.8	85	-0.01
85 c	Di-n-octyl phthalate	40.000	46.201	-15.5	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	73	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.327	9.2	61	0.00
88	Benzo(b)fluoranthene	40.000	43.701	-9.3	80	0.00
89	Benzo(k)fluoranthene	40.000	39.439	1.4	74	0.00
90 C	Benzo(a)pyrene	40.000	40.460	-1.2	74	0.00
91	Dibenzo(a,h)anthracene	40.000	36.867	7.8	62	0.00
92	Benzo(g,h,i)perylene	40.000	35.525	11.2	60	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072123\
Data File : BF134392.D
Acq On : 21 Jul 2023 09:59
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 21 13:26:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 14 22:51:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0