


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072320\
 Data File : BF120976.D
 Acq On : 23 Jul 2020 11:23
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 16 14:20:32 2020



AutoFind: Scans 1615, 1616, 1617; Background Corrected with Scan 1610

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	38810	PASS
68	69	0.00	2	1.5	544	PASS
69	198	0.00	100	32.2	35685	PASS
70	69	0.00	2	0.3	111	PASS
127	198	10	80	47.2	52322	PASS
197	198	0.00	2	0.7	742	PASS
198	198	100	100	100.0	110786	PASS
199	198	5	9	6.7	7464	PASS
275	198	10	60	27.3	30229	PASS
365	198	1	100	3.4	3760	PASS
441	198	0.01	100	12.6	13971	PASS
442	198	50	100	81.0	89757	PASS
443	442	15	24	18.8	16906	PASS