

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072423\
 Data File : BF134438.D
 Acq On : 24 Jul 2023 20:03
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 25 02:46:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 01:46:04 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	114	0.00
2	1,4-Dioxane	40.000	39.848	0.4	116	0.00
3	Pyridine	40.000	41.579	-3.9	112	0.00
4	n-Nitrosodimethylamine	40.000	42.572	-6.4	120	0.00
5 S	2-Fluorophenol	80.000	81.178	-1.5	111	0.00
6	Aniline	40.000	39.962	0.1	111	0.00
7 S	Phenol-d6	80.000	79.847	0.2	114	0.00
8	2-Chlorophenol	40.000	39.174	2.1	109	0.00
9	Benzaldehyde	40.000	37.187	7.0	112	0.00
10 C	Phenol	40.000	40.112	-0.3	113	0.00
11	bis(2-Chloroethyl)ether	40.000	39.014	2.5	112	0.00
12	1,3-Dichlorobenzene	40.000	39.595	1.0	109	0.00
13 C	1,4-Dichlorobenzene	40.000	39.275	1.8	110	0.00
14	1,2-Dichlorobenzene	40.000	39.315	1.7	113	0.00
15	Benzyl Alcohol	40.000	39.735	0.7	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.229	-0.6	115	0.00
17	2-Methylphenol	40.000	39.966	0.1	111	0.00
18	Hexachloroethane	40.000	38.159	4.6	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.499	3.8	109	0.00
20	3+4-Methylphenols	40.000	38.619	3.5	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	40.446	-1.1	109	0.00
23 S	Nitrobenzene-d5	80.000	83.267	-4.1	111	0.00
24	Nitrobenzene	40.000	41.286	-3.2	113	0.00
25	Isophorone	40.000	41.019	-2.5	111	0.00
26 C	2-Nitrophenol	40.000	42.422	-6.1	113	0.00
27	2,4-Dimethylphenol	40.000	41.299	-3.2	112	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.007	-2.5	112	0.00
29 C	2,4-Dichlorophenol	40.000	40.990	-2.5	109	0.00
30	1,2,4-Trichlorobenzene	40.000	40.650	-1.6	109	0.00
31	Naphthalene	40.000	40.309	-0.8	109	0.00
32	Benzoic acid	40.000	40.468	-1.2	104	0.00
33	4-Chloroaniline	40.000	40.884	-2.2	109	0.00
34 C	Hexachlorobutadiene	40.000	41.334	-3.3	110	0.00
35	Caprolactam	40.000	42.540	-6.3	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.319	-0.8	108	0.00
37	2-Methylnaphthalene	40.000	40.616	-1.5	109	0.00
38	1-Methylnaphthalene	40.000	39.739	0.7	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.824	-2.1	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.535	-1.3	105	0.00
42 S	2,4,6-Tribromophenol	80.000	81.685	-2.1	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.669	-1.7	96	0.00
44	2,4,5-Trichlorophenol	40.000	41.442	-3.6	97	0.00
45 S	2-Fluorobiphenyl	80.000	80.487	-0.6	96	0.00
46	1,1'-Biphenyl	40.000	40.554	-1.4	96	0.00
47	2-Chloronaphthalene	40.000	40.976	-2.4	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072423\
 Data File : BF134438.D
 Acq On : 24 Jul 2023 20:03
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 25 02:46:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 01:46:04 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.120	-2.8	95	0.00
49	Acenaphthylene	40.000	40.731	-1.8	96	0.00
50	Dimethylphthalate	40.000	40.913	-2.3	97	0.00
51	2,6-Dinitrotoluene	40.000	40.699	-1.7	95	0.00
52 C	Acenaphthene	40.000	40.777	-1.9	96	0.00
53	3-Nitroaniline	40.000	41.322	-3.3	97	0.00
54 P	2,4-Dinitrophenol	40.000	37.089	7.3	92	0.00
55	Dibenzofuran	40.000	40.788	-2.0	97	0.00
56 P	4-Nitrophenol	40.000	41.596	-4.0	98	0.00
57	2,4-Dinitrotoluene	40.000	41.340	-3.4	95	0.00
58	Fluorene	40.000	40.775	-1.9	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.332	4.2	90	0.00
60	Diethylphthalate	40.000	40.520	-1.3	97	0.00
61	4-Chlorophenyl-phenylether	40.000	40.686	-1.7	96	0.00
62	4-Nitroaniline	40.000	40.617	-1.5	94	0.00
63	Azobenzene	40.000	40.596	-1.5	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.947	-2.4	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.410	-1.0	96	0.00
67	4-Bromophenyl-phenylether	40.000	39.426	1.4	95	0.00
68	Hexachlorobenzene	40.000	39.807	0.5	96	0.00
69	Atrazine	40.000	41.029	-2.6	96	0.00
70 C	Pentachlorophenol	40.000	41.617	-4.0	98	0.00
71	Phenanthrene	40.000	40.446	-1.1	97	0.00
72	Anthracene	40.000	40.166	-0.4	95	0.00
73	Carbazole	40.000	40.157	-0.4	97	0.00
74	Di-n-butylphthalate	40.000	39.364	1.6	94	0.00
75 C	Fluoranthene	40.000	40.057	-0.1	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzidine	40.000	43.498	-8.7	103	0.00
78	Pyrene	40.000	39.728	0.7	97	0.00
79 S	Terphenyl-d14	80.000	77.078	3.7	94	0.00
80	Butylbenzylphthalate	40.000	40.037	-0.1	95	0.00
81	Benzo(a)anthracene	40.000	40.496	-1.2	100	0.00
82	3,3'-Dichlorobenzidine	40.000	40.402	-1.0	99	0.00
83	Chrysene	40.000	41.331	-3.3	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.219	-5.5	101	0.00
85 c	Di-n-octyl phthalate	40.000	45.350	-13.4	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.405	-1.0	99	-0.01
88	Benzo(b)fluoranthene	40.000	39.742	0.6	105	0.00
89	Benzo(k)fluoranthene	40.000	42.257	-5.6	104	0.00
90 C	Benzo(a)pyrene	40.000	41.363	-3.4	104	0.00
91	Dibenzo(a,h)anthracene	40.000	41.118	-2.8	101	0.00
92	Benzo(g,h,i)perylene	40.000	40.846	-2.1	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072423\
Data File : BF134438.D
Acq On : 24 Jul 2023 20:03
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 25 02:46:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 01:46:04 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0