

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072518\
 Data File : BF107679.D
 Acq On : 26 Jul 2018 2:20
 Operator : JU/SJ
 Sample : J4126-05DL 40X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-11DL

Manual Integrations
 APPROVED

Sohil
 7/26/2018 2:02:26 PM

Quant Time: Jul 26 04:46:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF072018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 20 16:29:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	127619	20.00	ng	-0.01
21) Naphthalene-d8	8.25	136	530687	20.00	ng	-0.01
38) Acenaphthene-d10	10.01	164	206432	20.00	ng	-0.01
63) Phenanthrene-d10	11.51	188	374021	20.00	ng	0.00
75) Chrysene-d12	14.15	240	339902	20.00	ng	-0.01
86) Perylene-d12	15.69	264	297361	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	2042	0.26	ng	0.02
7) Phenol-d6	6.66	99	8696	0.91	ng	0.05
23) Nitrobenzene-d5	7.55	82	4671	0.59	ng	0.00
41) 2,4,6-Tribromophenol	10.81	330	4130	1.76	ng	-0.01
44) 2-Fluorobiphenyl	0.00	172	0d	0.00	ng	
78) Terphenyl-d14	13.08	244	20263	1.53	ng	-0.02
Target Compounds						
31) Naphthalene	8.27	128	53738	2.303	ng	98
45) 1,1'-Biphenyl	9.42	154	60826	3.382	ng	94
48) Acenaphthylene	9.87	152	46986	2.276	ng	98
51) Acenaphthene	10.04	154	306205	23.668	ng	99
57) Fluorene	10.56	166	185872	13.679	ng	93
70) Phenanthrene	11.53	178	1316733	72.244	ng	99
71) Anthracene	11.58	178	402237	21.919	ng	99
74) Fluoranthene	12.72	202	805852	41.203	ng	96
77) Pyrene	12.95	202	1022959	43.999	ng	98
80) Benzo(a)anthracene	14.15	228	336507m	16.894	ng	
82) Chrysene	14.18	228	242170	12.656	ng	100
85) Indeno(1,2,3-cd)pyrene	17.27	276	79154	4.881	ng	# 90
87) Benzo(b)fluoranthene	15.24	252	216515m	13.304	ng	
88) Benzo(k)fluoranthene	15.26	252	69431m	4.355	ng	
89) Benzo(a)pyrene	15.62	252	206614m	13.879	ng	
91) Benzo(g,h,i)perylene	17.75	276	77880	6.136	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072518\
Data File : BF107679.D
Acq On : 26 Jul 2018 2:20
Operator : JU/SJ
Sample : J4126-05DL 40X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-11DL

Manual Integrations
APPROVED

Sohil
7/26/2018 2:02:26 PM

Quant Time: Jul 26 04:46:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF072018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 20 16:29:27 2018
Response via : Initial Calibration

